



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107483

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/31/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	01/31/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	01/31/2020	17:17
4	ICV	1	Water	NA	NA	20.1	7.00	01/31/2020	17:19
5	CCV1	1	Water	NA	NA	20.1	2.01	01/31/2020	17:20
6	L1016-19	1	Solid	20.03	20	21.0	8.86	01/31/2020	17:22
7	L1343-02	1	Solid	20.04	20	21.3	9.73	01/31/2020	17:25
8	L1343-02DUP	1	Solid	20.02	20	21.4	9.74	01/31/2020	17:26
9	CCV2	1	Water	NA	NA	20.2	12.02	01/31/2020	17:30

18107683

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1343 WorkList ID : 135621 Department : Wet-Chemistry Date : 01-31-2020 14:51:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2020	Solid	L1016-19	Corrosivity	Cool 4 deg C	PSEG05		SU-02-013120	01/31/2020	9045D
01/31/2020	Solid	L1343-02	Corrosivity	Cool 4 deg C	PSEG01	I51	NW-SP	01/31/2020	9045D

Date/Time 01-31-2020 16:00
Received by: JP
Relinquished by: JC

Date/Time 01-31-2020 17:35
Received by: L.C.
Relinquished by: JP