



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107485

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/31/2020	09:25
2	L1329-14		1	8	77.00	170.6	170.6	01/31/2020	09:50
3	L1329-14DUP		1	9	78.00	172.4	172.4	01/31/2020	10:20
4	L1329-20		1	7	61.00	141.8	141.8	01/31/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 107485

WorkList Name : fp13120-1 WorkList ID : 135587 Department : Wet-Chemistry Date : 01-31-2020 08:47:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2020	Solid	L1329-14 B	Flash Point	Cool 4 deg C	PSEG01	I43	40634	01/30/2020	1010A
02/04/2020	Solid	L1329-20 B	Flash Point	Cool 4 deg C	PSEG01	I43	40510	01/30/2020	1010A

Date/Time 01/31/20 08:55
 Received by: KS
 Relinquished by: JS

Date/Time 01/31/20 10:50
 Received by: JS
 Relinquished by: KS