



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 02/03/2020

Run Number: LB107491

BalanceID: WC SC-3

OvenID: WC OVEN-1

FilterID: 1827-047

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 02/03/2020 07:10 TEMP1 OUT: 104 °C 02/03/2020 08:10
TEMP2 IN: 104 °C 02/03/2020 08:40 TEMP2 OUT: 104 °C 02/03/2020 10:00
TEMP3 IN: 103 °C 02/03/2020 11:25 TEMP3 OUT: 103 °C 02/03/2020 13:00
TEMP4 IN: 103 °C 02/03/2020 13:30 TEMP4 OUT: 103 °C 02/03/2020 15:20

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB107491BL	LB107491BL	2.4169	2.4169	100	2.4169	2.4169	2.4169	0.0000	0
2	LB107491BS	LB107491BS	1.3257	1.3257	100	1.3804	1.3804	1.3804	0.0547	547
3	L1352-01	012820-DBRI-F-D-B	1.3723	1.3723	500	1.3949	1.3949	1.3949	0.0226	45.2
4	L1352-02	012820-DBRI-F-D-M	1.3739	1.3739	500	1.3977	1.3977	1.3977	0.0238	47.6
5	L1352-03	012820-DBRI-F-D-S	1.3757	1.3757	500	1.3897	1.3897	1.3897	0.0140	28
6	L1352-04	012820-DBRI-F-U-B	1.3821	1.3821	500	1.4139	1.4139	1.4139	0.0318	63.6
7	L1352-05	012820-DBRI-F-U-M	1.3812	1.3812	800	1.4113	1.4113	1.4113	0.0301	37.6
8	L1352-06	012820-DBRI-F-U-S	1.3767	1.3767	800	1.3974	1.3974	1.3974	0.0207	25.9
9	L1352-07	012820-TIPC-F-D-B	1.3810	1.3810	500	1.4095	1.4095	1.4095	0.0285	57
10	L1352-08	012820-TIPC-F-D-S	1.3776	1.3776	500	1.4127	1.4127	1.4127	0.0351	70.2
11	L1352-09	012820-TIPC-F-U-B	1.3809	1.3809	500	1.4015	1.4015	1.4015	0.0206	41.2
12	L1352-10	012820-TIPC-F-U-S	1.3853	1.3853	500	1.4069	1.4069	1.4069	0.0216	43.2
13	L1352-10DUP	012820-TIPC-F-U-SDUP	1.3812	1.3812	500	1.4038	1.4038	1.4038	0.0226	45.2
14	L1353-01	013120-TIPC-F-D-B	1.3767	1.3767	500	1.3896	1.3896	1.3896	0.0129	25.8
15	L1353-02	013120-TIPC-F-D-S	1.3726	1.3726	800	1.3908	1.3908	1.3908	0.0182	22.8



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 02/03/2020

Run Number: LB107491

BalanceID: WC SC-3

OvenID: WC OVEN-1

FilterID: 1827-047

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 02/03/2020 07:10 TEMP1 OUT: 104 °C 02/03/2020 08:10
TEMP2 IN: 104 °C 02/03/2020 08:40 TEMP2 OUT: 104 °C 02/03/2020 10:00
TEMP3 IN: 103 °C 02/03/2020 11:25 TEMP3 OUT: 103 °C 02/03/2020 13:00
TEMP4 IN: 103 °C 02/03/2020 13:30 TEMP4 OUT: 103 °C 02/03/2020 15:20

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	L1353-03	013120-TIPC-F-U-B	1.3800	1.3800	500	1.3934	1.3934	1.3934	0.0134	26.8
17	L1353-04	013120-TIPC-F-U-S	1.3689	1.3689	500	1.3816	1.3816	1.3816	0.0127	25.4
18	L1354-01	013020-TIPC-F-D-B	1.3719	1.3719	500	1.4237	1.4237	1.4237	0.0518	103.6
19	L1354-02	013020-TIPC-F-D-S	1.3741	1.3741	500	1.3964	1.3964	1.3964	0.0223	44.6
20	L1354-03	013020-TIPC-F-U-B	1.3699	1.3699	500	1.3951	1.3951	1.3951	0.0252	50.4
21	L1354-04	013020-TIPC-F-U-S	1.3766	1.3766	500	1.4030	1.4030	1.4030	0.0264	52.8
22	L1354-05	013020-JETT-F-D-B	1.3729	1.3729	800	1.3947	1.3947	1.3947	0.0218	27.3
23	L1354-06	013020-JETT-F-D-M	1.3713	1.3713	800	1.3906	1.3906	1.3906	0.0193	24.1
24	L1354-07	013020-JETT-F-D-S	1.3719	1.3719	800	1.3883	1.3883	1.3883	0.0164	20.5
25	L1354-08	013020-JETT-F-U-B	1.3832	1.3832	500	1.4287	1.4287	1.4287	0.0455	91
26	L1354-09	013020-JETT-F-U-M	1.3854	1.3854	800	1.4052	1.4052	1.4052	0.0198	24.8
27	L1354-10	013020-JETT-F-U-S	1.3765	1.3765	500	1.3901	1.3901	1.3901	0.0136	27.2
28	L1354-10DUP	013020-JETT-F-U-SDUP	1.3791	1.3791	500	1.3929	1.3929	1.3929	0.0138	27.6



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 02/03/2020

Run Number: LB107491

BalanceID: WC SC-3

OvenID: WC OVEN-1

FilterID: 1827-047

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 02/03/2020 07:10 TEMP1 OUT: 104 °C 02/03/2020 08:10
TEMP2 IN: 104 °C 02/03/2020 08:40 TEMP2 OUT: 104 °C 02/03/2020 10:00
TEMP3 IN: 103 °C 02/03/2020 11:25 TEMP3 OUT: 103 °C 02/03/2020 13:00
TEMP4 IN: 103 °C 02/03/2020 13:30 TEMP4 OUT: 103 °C 02/03/2020 15:20

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)
B = Empty Dish Weight (g)
C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)
D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

28 107491

WorkList Name : TSS L1352 WorkList ID : 135633 Department : Wet-Chemistry Date : 02-03-2020 07:57:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2020	Water	L1352-01	TSS	Cool 4 deg C	HDR103	L12	012820-DBRI-F-D-B	01/28/2020	SM2540 D
02/04/2020	Water	L1352-02	TSS	Cool 4 deg C	HDR103	L12	012820-DBRI-F-D-M	01/28/2020	SM2540 D
02/04/2020	Water	L1352-03	TSS	Cool 4 deg C	HDR103	L12	012820-DBRI-F-D-S	01/28/2020	SM2540 D
02/04/2020	Water	L1352-04	TSS	Cool 4 deg C	HDR103	L12	012820-DBRI-F-U-B	01/28/2020	SM2540 D
02/04/2020	Water	L1352-05	TSS	Cool 4 deg C	HDR103	L12	012820-DBRI-F-U-M	01/28/2020	SM2540 D
02/04/2020	Water	L1352-06	TSS	Cool 4 deg C	HDR103	L12	012820-DBRI-F-U-S	01/28/2020	SM2540 D
02/04/2020	Water	L1352-07	TSS	Cool 4 deg C	HDR103	L12	012820-TIPC-F-D-B	01/28/2020	SM2540 D
02/04/2020	Water	L1352-08	TSS	Cool 4 deg C	HDR103	L12	012820-TIPC-F-D-S	01/28/2020	SM2540 D
02/04/2020	Water	L1352-09	TSS	Cool 4 deg C	HDR103	L12	012820-TIPC-F-U-B	01/28/2020	SM2540 D
02/04/2020	Water	L1352-10	TSS	Cool 4 deg C	HDR103	L12	012820-TIPC-F-U-S	01/28/2020	SM2540 D
02/07/2020	Water	L1353-01	TSS	Cool 4 deg C	HDR103	L12	013120-TIPC-F-D-B	01/31/2020	SM2540 D
02/07/2020	Water	L1353-02	TSS	Cool 4 deg C	HDR103	L12	013120-TIPC-F-D-S	01/31/2020	SM2540 D
02/07/2020	Water	L1353-03	TSS	Cool 4 deg C	HDR103	L12	013120-TIPC-F-U-B	01/31/2020	SM2540 D
02/07/2020	Water	L1353-04	TSS	Cool 4 deg C	HDR103	L12	013120-TIPC-F-U-S	01/31/2020	SM2540 D
02/06/2020	Water	L1354-01	TSS	Cool 4 deg C	HDR103	L12	013020-TIPC-F-D-B	01/30/2020	SM2540 D
02/06/2020	Water	L1354-02	TSS	Cool 4 deg C	HDR103	L12	013020-TIPC-F-D-S	01/30/2020	SM2540 D
02/06/2020	Water	L1354-03	TSS	Cool 4 deg C	HDR103	L12	013020-TIPC-F-U-B	01/30/2020	SM2540 D
02/06/2020	Water	L1354-04	TSS	Cool 4 deg C	HDR103	L12	013020-TIPC-F-U-S	01/30/2020	SM2540 D
02/06/2020	Water	L1354-05	TSS	Cool 4 deg C	HDR103	L12	013020-JETT-F-D-B	01/30/2020	SM2540 D
02/06/2020	Water	L1354-06	TSS	Cool 4 deg C	HDR103	L12	013020-JETT-F-D-M	01/30/2020	SM2540 D
02/06/2020	Water	L1354-07	TSS	Cool 4 deg C	HDR103	L12	013020-JETT-F-D-S	01/30/2020	SM2540 D
02/06/2020	Water	L1354-08	TSS	Cool 4 deg C	HDR103	L12	013020-JETT-F-U-B	01/30/2020	SM2540 D

Date/Time 02/03/2020 10:00
 Received by: Jo
 Relinquished by: Jo

Date/Time 02/03/2020 12:40
 Received by: Jo
 Relinquished by: Jo

WORKLIST(Hardcopy Internal Chain)

WB1079) WB107491

WorkList Name : TSS L1352

WorkList ID : 135633

Department : Wet-Chemistry

Date : 02-03-2020 07:57:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2020	Water	L1354-09	TSS	Cool 4 deg C	HDR103	L12	013020-JETT-F-U-M	01/30/2020	SM2540 D
02/06/2020	Water	L1354-10	TSS	Cool 4 deg C	HDR103	L12	013020-JETT-F-U-S	01/30/2020	SM2540 D

Date/Time 02/03/2020 10:00
 Received by: JD
 Relinquished by: JD

Date/Time 02/03/2020 12:40
 Received by: JD
 Relinquished by: JD