



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107495

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/03/2020	11:15
2	CAL2	1	Water	NA	NA	20.1	7.00	02/03/2020	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	02/03/2020	11:17
4	ICV	1	Water	NA	NA	20.2	7.00	02/03/2020	11:20
5	CCV1	1	Water	NA	NA	20.1	2.01	02/03/2020	11:22
6	L1347-02	1	Solid	20.03	20	22.1	7.96	02/03/2020	11:25
7	L1347-02DUP	1	Solid	20.04	20	22.2	7.97	02/03/2020	11:26
8	CCV2	1	Water	NA	NA	20.1	12.02	02/03/2020	11:30

213107495

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1347

WorkList ID : 135651

Department : Wet-Chemistry

Date : 02-03-2020 11:08:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2020	Solid	L1347-02 -17	pH	Cool 4 deg C	NEWP02	J51	COMP-01	01/31/2020	9045D

Date/Time 02-03-2020 11:10
Received by: JN
Relinquished by: A

Date/Time 02-03-2020 12:20
Received by: [Signature]
Relinquished by: JN