

LB107508


7470A

Instrument 2D: CV1

Linear

 μ Abs.:

39714



A= 0.0000e+000

B= 2.5460e-004

C= -1.0107e-001

Rho= 0.9998472

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	502	0.000	502				
0.2	0.200	0.226	0.026	1286	0.0 %	1286				
2.5	2.500	2.499	-0.001	10212	0.0 %	10212				
5.0	5.000	4.865	-0.135	19506	0.0 %	19506				
7.5	7.500	7.573	0.073	30140	0.0 %	30140				
10.0	10.000	10.010	0.010	39714	0.0 %	39714				

LB107508 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Feb 2020 15:34:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	502	-	0.0000	7470A	PPB	03 Feb 2020 15:45:26	S	Std
0.2 - 1	50.2	1286	-	0.2000	7470A	PPB	03 Feb 2020 15:47:34	S	Std
2.5 - 1	51.5	10212	-	2.5000	7470A	PPB	03 Feb 2020 15:49:41	S	Std
5.0 - 1	55.0	19506	-	5.0000	7470A	PPB	03 Feb 2020 15:51:48	S	Std
7.5 - 1	57.5	30140	-	7.5000	7470A	PPB	03 Feb 2020 15:53:55	S	Std
10.0 - 1	50.0	39714	-	10.0000	7470A	PPB	03 Feb 2020 15:56:02	S	Std
ICV96 - 1	ICV96	16007	3.9743	-	7470A	PPB	03 Feb 2020 16:01:44	U	SMPL
ICB96 - 1	ICB96	299	-0.0249	-	7470A	PPB	03 Feb 2020 16:03:49	U	SMPL
CCV62 - 1	CCV62	19550	4.8764	-	7470A	PPB	03 Feb 2020 16:05:55	U	SMPL
CCB62 - 1	CCB62	277	-0.0305	-	7470A	PPB	03 Feb 2020 16:08:00	U	SMPL
CRA - 1	CRA	1221	0.2098	-	7470A	PPB	03 Feb 2020 16:10:06	U	SMPL
HighStd - 1	HighStd	39095	9.8526	-	7470A	PPB	03 Feb 2020 16:12:11	U	SMPL
ChkStd - 1	ChkStd	27807	6.9786	-	7470A	PPB	03 Feb 2020 16:14:17	U	SMPL
PB126529BL - 1	PBW	269	-0.0326	-	7470A	PPB	03 Feb 2020 16:16:24	U	SMPL
PB126529BS - 1	LCSW	15985	3.9687	-	7470A	PPB	03 Feb 2020 16:18:30	U	SMPL
L1329-16 - 1	40230	61	-0.0855	-	7470A	PPB	03 Feb 2020 16:20:36	U	SMPL
L1329-16DUP - 1	40230DUP	26	-0.0945	-	7470A	PPB	03 Feb 2020 16:22:43	U	SMPL
L1329-16MS - 1	40230MS	15444	3.8310	-	7470A	PPB	03 Feb 2020 16:24:50	U	SMPL
L1329-16MSD - 1	40230MSD	16195	4.0222	-	7470A	PPB	03 Feb 2020 16:26:58	U	SMPL
L1329-17 - 1	40098	410	0.0033	-	7470A	PPB	03 Feb 2020 16:29:05	U	SMPL
CCV63 - 1	CCV63	19750	4.9273	-	7470A	PPB	03 Feb 2020 16:31:11	U	SMPL
CCB63 - 1	CCB63	300	-0.0247	-	7470A	PPB	03 Feb 2020 16:33:16	U	SMPL
L1329-18 - 1	COMP-11	531	0.0341	-	7470A	PPB	03 Feb 2020 16:35:22	U	SMPL
L1329-19 - 1	COMP-5	357	-0.0102	-	7470A	PPB	03 Feb 2020 16:37:28	U	SMPL
PB126530BL - 1	PBW	205	-0.0489	-	7470A	PPB	03 Feb 2020 16:39:34	U	SMPL
PB126530BS - 1	LCSW	15855	3.9356	-	7470A	PPB	03 Feb 2020 16:41:40	U	SMPL
L1320-04 - 1	ETGI-367	148	-0.0634	-	7470A	PPB	03 Feb 2020 16:43:47	U	SMPL
L1325-02 - 1	346	132	-0.0675	-	7470A	PPB	03 Feb 2020 16:45:54	U	SMPL
L1325-04 - 1	200029	154	-0.0619	-	7470A	PPB	03 Feb 2020 16:48:00	U	SMPL
L1325-06 - 1	VNJ-232	136	-0.0664	-	7470A	PPB	03 Feb 2020 16:50:08	U	SMPL
L1325-08 - 1	72-11976	132	-0.0675	-	7470A	PPB	03 Feb 2020 16:52:16	U	SMPL
L1325-10 - 1	200059	143	-0.0647	-	7470A	PPB	03 Feb 2020 16:54:22	U	SMPL
CCV64 - 1	CCV64	19485	4.8598	-	7470A	PPB	03 Feb 2020 16:56:28	U	SMPL
CCB64 - 1	CCB64	271	-0.0321	-	7470A	PPB	03 Feb 2020 16:58:34	U	SMPL
L1333-06 - 1	TP-1-2-COMP	127	-0.0687	-	7470A	PPB	03 Feb 2020 17:00:40	U	SMPL
L1333-06DUP - 1	TP-1-2-COMPDUP	138	-0.0659	-	7470A	PPB	03 Feb 2020 17:02:46	U	SMPL
L1333-06MS - 1	TP-1-2-COMPMS	15978	3.9670	-	7470A	PPB	03 Feb 2020 17:04:53	U	SMPL
L1333-06MSD - 1	TP-1-2-COMPMSD	15715	3.9000	-	7470A	PPB	03 Feb 2020 17:06:59	U	SMPL
L1329-16LX5 - 1		-10	-0.1036	-	7470A	PPB	03 Feb 2020 17:09:05	U	SMPL
L1333-06LX5 - 1		22	-0.0955	-	7470A	PPB	03 Feb 2020 17:11:12	U	SMPL
PB126530TB - 1	126486TB	334	-0.0160	-	7470A	PPB	03 Feb 2020 17:13:19	U	SMPL
CCV65 - 1	CCV65	19278	4.8071	-	7470A	PPB	03 Feb 2020 17:15:26	U	SMPL
CCB65 - 1	CCB65	282	-0.0293	-	7470A	PPB	03 Feb 2020 17:17:33	U	SMPL