



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107513

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/03/2020	10:35
2	L1329-16		1	9	100.00	>212.0	>212.0	02/03/2020	10:55
3	L1329-16DUP		1	9	100.00	>212.0	>212.0	02/03/2020	11:25
4	L1329-17		1	8	100.00	>212.0	>212.0	02/03/2020	11:50
5	L1329-18		1	7	100.00	>212.0	>212.0	02/03/2020	12:20
6	L1329-19		1	8	100.00	>212.0	>212.0	02/03/2020	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB107513

WorkList Name : fp13120 WorkList ID : 135586 Department : Wet-Chemistry Date : 01-31-2020 08:45:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2020	Water	L1329-16	Flash Point	Cool 4 deg C	PSEG01	I43	40230	01/30/2020	1010A
02/04/2020	Water	L1329-17	Flash Point	Cool 4 deg C	PSEG01	I43	40098	01/30/2020	1010A
02/04/2020	Water	L1329-18	Flash Point	Cool 4 deg C	PSEG01	I43	COMP-11	01/30/2020	1010A
02/04/2020	Water	L1329-19	Flash Point	Cool 4 deg C	PSEG01	I43	COMP-5	01/30/2020	1010A

Date/Time 02/03/20 10:05
 Received by: KS
 Relinquished by: KS

Date/Time 02/03/20 12:50
 Received by: Pe
 Relinquished by: KS