



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107518

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/03/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	02/03/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	02/03/2020	17:19
4	ICV	1	Water	NA	NA	20.1	7.00	02/03/2020	17:20
5	CCV1	1	Water	NA	NA	20.1	2.01	02/03/2020	17:22
6	L1377-01	1	Water	NA	NA	16.6	7.24	02/03/2020	17:25
7	L1377-01DUP	1	Water	NA	NA	16.7	7.25	02/03/2020	17:26
8	L1377-05	1	Water	NA	NA	16.5	7.17	02/03/2020	17:30
9	CCV2	1	Water	NA	NA	20.1	12.02	02/03/2020	17:32

18107518

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1377

WorkList ID : 135668

Department : Wet-Chemistry

Date : 02-03-2020 17:02:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/03/2020	Water	L1377-01-04	pH	Cool 4 deg C	CDMS01	L21	BW-2	02/03/2020	9040C
02/03/2020	Water	L1377-05-04	pH	Cool 4 deg C	CDMS01	L21	BW-209	02/03/2020	9040C

Date/Time 02/03/2020 17:10
 Received by: Ja
 Relinquished by: KS

Date/Time 02-03-2020 18:20
 Received by: KS
 Relinquished by: Ja