

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 02/05/2020

OVEN IN TEMP1: 98 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 02/04/2020 16:20

DATE & TIME OUT1: 02/05/2020 07:10

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 02/05/2020 08:00

DATE & TIME OUT2: 02/05/2020 09:35

Run Number: LB107521

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB107521BL	LB107521BL	96.5137	96.5137	100	96.5137	96.5137	96.5137	0.0000	0
2	L1377-01	BW-2	85.4483	85.4483	100	85.4785	85.4785	85.4785	0.0302	302
3	L1377-01DUP	BW-2DUP	85.0849	85.0849	100	85.1159	85.1159	85.1159	0.0310	310
4	L1377-05	BW-209	90.7556	90.7556	100	90.7904	90.7904	90.7904	0.0348	348

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

2B 107521

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS L1377 WorkList ID : 135688 Department : Wet-Chemistry Date : 02-04-2020 09:33:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/10/2020	Water	L1377-01	TS	Cool 4 deg C	CDMS01	L21	BW-2	02/03/2020	SM2540 B
02/10/2020	Water	L1377-05	TS	Cool 4 deg C	CDMS01	L21	BW-209	02/03/2020	SM2540 B

Date/Time 02-04-2020 12/1/20
Received by: JA
Relinquished by: KS

Date/Time 02/04/2020 16.20
Received by: KS
Relinquished by: JP