

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB107524BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP81750
calibration std. phosphate 0.5 ppm	WP81749
calibration std. phosphate 0.3 ppm	WP81748
calibration std. phosphate 0.1 ppm	WP81747
calibration std. phosphate 0.05 ppm	WP81746
calibration std. 0 ppm	WP81745
phosphate ICV std.	WP81752
5N sulfuric acid	WP79628
Combined reagent	WP81801
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP81711
Sodium hydroxide, 1N	WP77314
Phosphate LOD std, 0.025ppm	WP81802

Intercept: -0.0006Slope: 0.6501Regression: 0.999825

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	02/04/2020	11:48
2	CAL2	0.05	1	50	50	0.033	02/04/2020	11:49
3	CAL3	0.10	1	50	50	0.066	02/04/2020	11:50
4	CAL4	0.30	1	50	50	0.186	02/04/2020	11:50
5	CAL5	0.50	1	50	50	0.330	02/04/2020	11:51
6	CAL6	1.00	1	50	50	0.649	02/04/2020	11:51



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB107524

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.499	02/04/2020	11:52
2	ICB		1	50	50	0.000	0.001	02/04/2020	11:53
3	CCV1	0.50	1	50	50	0.329	0.507	02/04/2020	11:53
4	CCB1		1	50	50	0.000	0.001	02/04/2020	11:53
5	PB126566BL		1	1.00	50	0.000	0.001	02/04/2020	11:54
6	PB126566BS	0.50	1	1.00	50	0.320	0.493	02/04/2020	11:54
7	L1161-03		1	1.01	50	0.017	0.027	02/04/2020	11:55
8	CCV2	0.50	1	50	50	0.324	0.499	02/04/2020	11:55
9	CCB2		1	50	50	0.000	0.001	02/04/2020	11:56