



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107531

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/04/2020	13:05
2	CAL2	1	Water	NA	NA	20.1	7.00	02/04/2020	13:06
3	CAL3	1	Water	NA	NA	20.1	10.04	02/04/2020	13:08
4	ICV	1	Water	NA	NA	20.2	7.01	02/04/2020	13:10
5	CCV1	1	Water	NA	NA	20.2	2.01	02/04/2020	13:11
6	L1387-01	1	Water	NA	NA	15.3	6.92	02/04/2020	13:12
7	L1387-01DUP	1	Water	NA	NA	15.4	6.93	02/04/2020	13:13
8	L1388-01	1	Water	NA	NA	14.8	12.22	02/04/2020	13:20
9	CCV2	1	Water	NA	NA	20.1	12.02	02/04/2020	13:22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1378

WorkList ID : 135699

Department : Wet-Chemistry

Date : 02-04-2020 12:53:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2020	Water	L1387-01	pH	Cool 4 deg C	PSEG01	I52	F06031	02/04/2020	9040C
02/04/2020	Water	L1388-01	pH	Cool 4 deg C	PSEG01	I52	TOTE-COMP	02/04/2020	9040C

Date/Time

Received by:

Relinquished by:

02/04/2020 13:00

SP

DK

Date/Time

Received by:

Relinquished by:

02/04/2020 14:00

DK

IN