

LB 107534


ISM02.4

INSTRUMENT ID: CY1

Linear

 μ Abs.:

51057



A= 0.0000e+000

B= 1.9767e-004

C= -7.0666e-002

Rho= 0.9999867

Accept=Accepted

Slope
Intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% (1)
0.00	0.000	0.027	0.027	494	0.000	494					11
0.05	0.050										8
0.20	0.200	0.184	-0.016	1287	0.0 %	1287					0
2.50	2.500	2.496	-0.004	12985	0.0 %	12985					0
5.00	5.000	4.997	-0.003	25637	0.0 %	25637					0
7.50	7.500	7.474	-0.026	38167	0.0 %	38167					0
10.0	10.000	10.022	0.022	51057	0.0 %	51057					0

LB107534
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.00 - 1	S0.0	494	-	0	ISM02.4	PPB	2/4/2020 14:13	S	Std
0.20 - 1	S0.2	1287	-	0.2	ISM02.4	PPB	2/4/2020 14:15	S	Std
2.50 - 1	S2.5	12985	-	2.5	ISM02.4	PPB	2/4/2020 14:18	S	Std
5.00 - 1	S5.0	25637	-	5	ISM02.4	PPB	2/4/2020 14:20	S	Std
7.50 - 1	S7.5	38167	-	7.5	ISM02.4	PPB	2/4/2020 14:22	S	Std
10.0 - 1	S10.0	51057	-	10	ISM02.4	PPB	2/4/2020 14:24	S	Std
ICV - 1	ICV	20985	4.0775 -		ISM02.4	PPB	2/4/2020 14:39	U	SMPL
ICB - 1	ICB	445	0.0173 -		ISM02.4	PPB	2/4/2020 14:41	U	SMPL
CCV066 - 1	CCV066	25313	4.9331 -		ISM02.4	PPB	2/4/2020 14:43	U	SMPL
CCB066 - 1	CCB066	418	0.012 -		ISM02.4	PPB	2/4/2020 14:46	U	SMPL
PB126585BL - 1	PBW585	455	0.0193 -		ISM02.4	PPB	2/4/2020 14:48	U	SMPL
L1341-01 - 1	MBOAG5	1343	0.1948 -		ISM02.4	PPB	2/4/2020 14:50	U	SMPL
L1341-02 - 1	MBOAG6	17715	3.4311 -		ISM02.4	PPB	2/4/2020 14:52	U	SMPL
L1341-03 - 1	MBOAG7	1261	0.1786 -		ISM02.4	PPB	2/4/2020 14:55	U	SMPL
L1341-04 - 1	MBOAG8	11644	2.2311 -		ISM02.4	PPB	2/4/2020 14:57	U	SMPL
L1341-05 - 1	MBOAG9	795	0.0865 -		ISM02.4	PPB	2/4/2020 14:59	U	SMPL
L1341-06 - 1	MBOAH0	1564	0.2385 -		ISM02.4	PPB	2/4/2020 15:01	U	SMPL
L1341-07 - 1	MBOAH1	1541	0.234 -		ISM02.4	PPB	2/4/2020 15:04	U	SMPL
L1341-08 - 1	MBOAH2	966	0.1203 -		ISM02.4	PPB	2/4/2020 15:06	U	SMPL
L1341-09 - 1	MBOAH3	693	0.0663 -		ISM02.4	PPB	2/4/2020 15:08	U	SMPL
L1341-10 - 1	MBOAH4	2095	0.3435 -		ISM02.4	PPB	2/4/2020 15:10	U	SMPL
L1341-11 - 1	MBOAH5	1298	0.1859 -		ISM02.4	PPB	2/4/2020 15:13	U	SMPL
L1341-12 - 1	MBOAH3D	713	0.0703 -		ISM02.4	PPB	2/4/2020 15:17	U	SMPL
L1341-13 - 1	MBOAH3S	13449	2.5879 -		ISM02.4	PPB	2/4/2020 15:20	U	SMPL
CCV067 - 1	CCV067	23926	4.6589 -		ISM02.4	PPB	2/4/2020 15:22	U	SMPL
CCB067 - 1	CCB067	418	0.012 -		ISM02.4	PPB	2/4/2020 15:24	U	SMPL