



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107542

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/04/2020	13:20
2	L1377-01		1	9	100.00	>212.0	>212.0	02/04/2020	13:45
3	L1377-01DUP		1	9	100.00	>212.0	>212.0	02/04/2020	14:15
4	L1377-05		1	8	100.00	>212.0	>212.0	02/04/2020	14:40
5	L1387-02		1	7	100.00	>212.0	>212.0	02/04/2020	15:10
6	L1388-02		1	8	100.00	>212.0	>212.0	02/04/2020	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 107542

WorkList Name : fp2420 WorkList ID : 135714 Department : Wet-Chemistry Date : 02-04-2020 16:40:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/17/2020	Water	L1377-01 O	Flash Point	Cool 4 deg C	CDMS01	L21	BW-2	02/03/2020	1010A
02/17/2020	Water	L1377-05 I	Flash Point	Cool 4 deg C	CDMS01	L21	BW-209	02/03/2020	1010A
02/07/2020	Water	L1387-02 C	Flash Point	Cool 4 deg C	PSEG01	I52	F06031	02/04/2020	1010A
02/07/2020	Water	L1388-02 C	Flash Point	Cool 4 deg C	PSEG01	I52	TOTE-COMP	02/04/2020	1010A

Date/Time 02/04/20 12:45
 Received by: KS
 Relinquished by: DK

Date/Time 02/04/20 15:40
 Received by: DK
 Relinquished by: KS