



## Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107550

Slope : 99.3

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2628         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 02/05/2020 | 09:05     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.01        | 02/05/2020 | 09:06     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 02/05/2020 | 09:07     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 02/05/2020 | 09:10     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.01        | 02/05/2020 | 09:11     |
| 6   | L1396-01    | 1  | Water  | NA          | NA          | 13.5             | 6.56        | 02/05/2020 | 09:12     |
| 7   | L1396-01DUP | 1  | Water  | NA          | NA          | 13.6             | 6.57        | 02/05/2020 | 09:13     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 02/05/2020 | 09:15     |

13.10.7550

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1396

WorkList ID : 135728

Department : Wet-Chemistry

Date : 02-05-2020 08:51:12

| Due Date   | Matrix | Sample     | Test | Preservative | Customer | Storage Location | Customer Sample    | Collect Date | Method      |
|------------|--------|------------|------|--------------|----------|------------------|--------------------|--------------|-------------|
| 02/04/2020 | Water  | L1396-01 ✓ | pH   | Cool 4 deg C | APT101   | L11              | WGGM-20-WW-TSF-007 | 02/04/2020   | SM 4500-H B |

Date/Time 02/05/2020 9:00  
Received by: fg  
Relinquished by: fg

Date/Time 02/05/2020 10:00  
Received by: fg  
Relinquished by: fg