

LB 107556


7470A

INSTRUMENT ID: CYL

Linear

 μ Abs.:

42525



A= 0.0000e+000

B= 2.3868e-004

C= -8.8759e-002

Rho= 0.9997947

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	278	0.000	278				
0.2	0.200	0.170	-0.030	1083	0.0 %	1083				
2.5	2.500	2.619	0.119	11343	0.0 %	11343				
5.0	5.000	4.993	-0.007	21291	0.0 %	21291				
7.5	7.500	7.380	-0.120	31294	0.0 %	31294				
10.0	10.000	10.061	0.061	42525	0.0 %	42525				

LB107556 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 05 Feb 2020 11:39:57					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	278	-	0.0000	7470A	PPB	05 Feb 2020 11:44:27	S	Std
0.2 - 1	50.2	1083	-	0.2000	7470A	PPB	05 Feb 2020 11:46:35	S	Std
2.5 - 1	52.5	11343	-	2.5000	7470A	PPB	05 Feb 2020 11:48:43	S	Std
5.0 - 1	55	21291	-	5.0000	7470A	PPB	05 Feb 2020 11:50:51	S	Std
7.5 - 1	57.5	31294	-	7.5000	7470A	PPB	05 Feb 2020 11:52:57	S	Std
10.0 - 1	510	42525	-	10.0000	7470A	PPB	05 Feb 2020 11:59:23	S	Std
ICV01 - 1	ICV01	16553	3.8620	-	7470A	PPB	05 Feb 2020 12:05:23	U	SMPL
ICB01 - 1	ICB01	264	-0.0257	-	7470A	PPB	05 Feb 2020 12:07:28	U	SMPL
CCV68 - 1	CCV68	19827	4.6435	-	7470A	PPB	05 Feb 2020 12:09:34	U	SMPL
CCB68 - 1	CCB68	276	-0.0229	-	7470A	PPB	05 Feb 2020 12:11:39	U	SMPL
CRA - 1	CRA	1030	0.1571	-	7470A	PPB	05 Feb 2020 12:13:45	U	SMPL
HighStd - 1	HighStd	42418	10.0354	-	7470A	PPB	05 Feb 2020 12:15:50	U	SMPL
ChkStd - 1	ChkStd	29192	6.8787	-	7470A	PPB	05 Feb 2020 12:17:56	U	SMPL
PB126595BL - 1	PBW	346	-0.0062	-	7470A	PPB	05 Feb 2020 12:20:04	U	SMPL
PB126595BS - 1	LCSW	16371	3.8186	-	7470A	PPB	05 Feb 2020 12:24:38	U	SMPL
L1016-19 - 1	SU-02-013120	634	0.0626	-	7470A	PPB	05 Feb 2020 12:26:45	U	SMPL
L1016-19DUP - 1	SU-02-013120DUP	542	0.0406	-	7470A	PPB	05 Feb 2020 12:28:52	U	SMPL
L1016-19MS - 1	SU-02-013120MS	14837	3.4525	-	7470A	PPB	05 Feb 2020 12:31:00	U	SMPL
L1016-19MSD - 1	SU-02-013120MSD	14568	3.3883	-	7470A	PPB	05 Feb 2020 12:33:08	U	SMPL
L1343-02 - 1	NW-SP	317	-0.0131	-	7470A	PPB	05 Feb 2020 12:35:14	U	SMPL
CCV69 - 1	CCV69	20234	4.7406	-	7470A	PPB	05 Feb 2020 12:37:20	U	SMPL
CCB69 - 1	CCB69	258	-0.0272	-	7470A	PPB	05 Feb 2020 12:39:26	U	SMPL
L1351-04 - 1	TAL-1	356	-0.0038	-	7470A	PPB	05 Feb 2020 12:41:32	U	SMPL
L1351-05 - 1	TAL-2	347	-0.0059	-	7470A	PPB	05 Feb 2020 12:43:38	U	SMPL
L1351-06 - 1	TAL-3	342	-0.0071	-	7470A	PPB	05 Feb 2020 12:45:44	U	SMPL
PB126596BL - 1	PBW	323	-0.0117	-	7470A	PPB	05 Feb 2020 12:47:50	U	SMPL
PB126596BS - 1	LCSW	16118	3.7582	-	7470A	PPB	05 Feb 2020 12:49:56	U	SMPL
L1363-05 - 1	HR-06-020320	320	-0.0124	-	7470A	PPB	05 Feb 2020 12:52:04	U	SMPL
L1363-05DUP - 1	HR-06-020320DUP	316	-0.0133	-	7470A	PPB	05 Feb 2020 12:54:11	U	SMPL
L1363-05MS - 1	HR-06-020320MS	13972	3.2460	-	7470A	PPB	05 Feb 2020 12:56:18	U	SMPL
L1363-05MSD - 1	HR-06-020320MSD	15398	3.5864	-	7470A	PPB	05 Feb 2020 13:01:11	U	SMPL
L1368-06 - 1	EO-01-020320	242	-0.0310	-	7470A	PPB	05 Feb 2020 13:03:16	U	SMPL
CCV70 - 1	CCV70	20108	4.7105	-	7470A	PPB	05 Feb 2020 13:05:22	U	SMPL
CCB70 - 1	CCB70	284	-0.0210	-	7470A	PPB	05 Feb 2020 13:07:28	U	SMPL
L1376-01 - 1	02-03-A	263	-0.0260	-	7470A	PPB	05 Feb 2020 13:09:34	U	SMPL
L1376-02 - 1	02-03-B	268	-0.0248	-	7470A	PPB	05 Feb 2020 13:11:41	U	SMPL
L1376-03 - 1	02-03-C	243	-0.0308	-	7470A	PPB	05 Feb 2020 13:13:47	U	SMPL
L1376-04 - 1	OILY-DEBRIS	271	-0.0241	-	7470A	PPB	05 Feb 2020 13:15:53	U	SMPL
L1383-04 - 1	TP-A-1	274	-0.0234	-	7470A	PPB	05 Feb 2020 13:18:00	U	SMPL
L1383-08 - 1	TP-B-1	243	-0.0308	-	7470A	PPB	05 Feb 2020 13:20:07	U	SMPL
L1383-12 - 1	TP-C-1	254	-0.0281	-	7470A	PPB	05 Feb 2020 13:22:14	U	SMPL
PB126597BL - 1	PBW	370	-0.0004	-	7470A	PPB	05 Feb 2020 13:24:21	U	SMPL
PB126597BS - 1	LCSW	16389	3.8229	-	7470A	PPB	05 Feb 2020 13:26:28	U	SMPL
L1387-01 - 1	F06031	1704	0.3179	-	7470A	PPB	05 Feb 2020 13:28:34	U	SMPL
CCV71 - 1	CCV71	21693	5.0888	-	7470A	PPB	05 Feb 2020 13:30:41	U	SMPL
CCB71 - 1	CCB71	352	-0.0047	-	7470A	PPB	05 Feb 2020 13:32:47	U	SMPL
L1387-01DUP - 1	F06031DUP	1533	0.2771	-	7470A	PPB	05 Feb 2020 13:38:02	U	SMPL
L1387-01MS - 1	F06031MS	18652	4.3630	-	7470A	PPB	05 Feb 2020 13:40:08	U	SMPL
L1387-01MSD - 1	F06031MSD	19307	4.5193	-	7470A	PPB	05 Feb 2020 13:42:14	U	SMPL
L1388-01 - 1	TOTE-COMP	123	-0.0594	-	7470A	PPB	05 Feb 2020 13:44:20	U	SMPL
L1016-19LX5 - 1		457	0.0203	-	7470A	PPB	05 Feb 2020 13:46:27	U	SMPL
L1016-19A - 1		17720	4.1406	-	7470A	PPB	05 Feb 2020 13:48:34	U	SMPL
L1363-05LX5 - 1		409	0.0089	-	7470A	PPB	05 Feb 2020 13:50:41	U	SMPL
L1363-05A - 1		16851	3.9332	-	7470A	PPB	05 Feb 2020 13:52:49	U	SMPL
L1387-01LX5 - 1		642	0.0645	-	7470A	PPB	05 Feb 2020 13:54:56	U	SMPL
L1387-01A - 1		20826	4.8819	-	7470A	PPB	05 Feb 2020 13:57:02	U	SMPL
CCV72 - 1	CCV72	21737	5.0993	-	7470A	PPB	05 Feb 2020 13:59:09	U	SMPL
CCB72 - 1	CCB72	389	0.0041	-	7470A	PPB	05 Feb 2020 14:01:15	U	SMPL
PB126595TB - 1	PB126525TB	162	-0.0501	-	7470A	PPB	05 Feb 2020 14:03:21	U	SMPL
PB126596TB - 1	PB126558TB	157	-0.0513	-	7470A	PPB	05 Feb 2020 14:05:27	U	SMPL
CCV73 - 1	CCV73	21134	4.9554	-	7470A	PPB	05 Feb 2020 14:07:33	U	SMPL
CCB73 - 1	CCB73	368	-0.0009	-	7470A	PPB	05 Feb 2020 14:09:39	U	SMPL