



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107570

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	02/05/2020	17:05
2	CAL2	1	Water	NA	NA	20.1	7.01	02/05/2020	17:06
3	CAL3	1	Water	NA	NA	20.1	10.03	02/05/2020	17:07
4	ICV	1	Water	NA	NA	20.2	7.00	02/05/2020	17:10
5	CCV1	1	Water	NA	NA	20.1	2.01	02/05/2020	17:11
6	L1365-08	1	Solid	20.03	20	19.8	8.43	02/05/2020	17:12
7	L1365-08DUP	1	Solid	20.04	20	19.9	8.44	02/05/2020	17:13
8	L1394-09	1	Solid	20.05	20	20.3	8.98	02/05/2020	17:14
9	CCV2	1	Water	NA	NA	20.2	12.02	02/05/2020	17:15

LB 107570

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L1394 WorkList ID : 135754 Department : Wet-Chemistry Date : 02-05-2020 15:49:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2020	Solid	L1365-08	Corrosivity	Cool 4 deg C	PSEG05	I53	CL-01-020520	02/05/2020	9045D
01/24/2020	Solid	L1394-09	Corrosivity	Cool 4 deg C	DIST01	L41	MNR-WHITE-PLAINS-COMP	01/24/2020	9045D

Date/Time 02-05-2020 16:05
Received by: Jg
Relinquished by: J.C

Date/Time 02-05-2020 17:45
Received by: J.C
Relinquished by: Jg