



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB107587

Reagent/Standard	Lot/Log #
Residual chlorine std, and CCV , 0.4PPM	WP81892
Residual chlorine LOD, 0.05PPM	WP81889
Residual chlorine ICV-LCS, 0.4PPM	WP81888
Chlorine Calibration std, 0.1ppm	WP81883
Chlorine Calibration std, 0.2ppm	WP81884
Chlorine Calibration std, 0.8ppm	WP81886
Chlorine Calibration std, 0.0ppm	WP81882
Chlorine Calibration std, 1.6ppm	WP81887
Total Chlorine Powder Pillows	W2249

Intercept: 0.0026

Slope: 1.0241

Regression: 0.999973

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	02/06/2020	11:46
2	CAL2	0.1	1	0.000	0.100	0.100	02/06/2020	11:47
3	CAL3	0.2	1	0.000	0.210	0.210	02/06/2020	11:48
4	CAL4	0.4	1	0.000	0.420	0.420	02/06/2020	11:49
5	CAL5	0.8	1	0.000	0.820	0.820	02/06/2020	11:50
6	CAL6	1.6	1	0.000	1.640	1.640	02/06/2020	11:51



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB107587

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3780	02/06/2020	11:52
2	ICB				1	0.0000	0.0000	0.0000	-0.0030	02/06/2020	11:53
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.3880	02/06/2020	11:54
4	CCB1				1	0.0000	0.0000	0.0000	-0.0030	02/06/2020	11:55
5	LB107587BL	50	50		1	0.0000	0.0000	0.0000	-0.0030	02/06/2020	11:56
6	LB107587BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3880	02/06/2020	11:57
7	L1161-09	50	50		1	0.0000	0.0400	0.0400	0.0370	02/06/2020	11:58
8	CCV2			0.4	1	0.0000	0.4200	0.4200	0.4080	02/06/2020	11:59
9	CCB2				1	0.0000	0.0000	0.0000	-0.0030	02/06/2020	12:00

WORKLIST(Hardcopy Internal Chain)

LB107587

WorkList Name : RCLMDL020620 WorkList ID : 135778 Department : Wet-Chemistry Date : 02-06-2020 11:10:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2020	Water	L1161-09	Residual Chlorine	Cool 4 deg C	CHEM02	QA Office	MDL-WATER-03-QT1-2020	01/16/2020	SM4500 Cl G

Date/Time _____
Received by: _____
Relinquished by: _____

Date/Time _____
Received by: _____
Relinquished by: _____