



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107599

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/06/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	7.00	02/06/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/06/2020	16:12
4	ICV	1	Water	NA	NA	20.2	7.00	02/06/2020	16:15
5	CCV1	1	Water	NA	NA	20.1	2.01	02/06/2020	16:16
6	L1419-02	1	Water	NA	NA	14.4	5.83	02/06/2020	16:19
7	L1419-04	1	Water	NA	NA	14.3	7.20	02/06/2020	16:20
8	L1419-05	1	Water	NA	NA	14.7	7.35	02/06/2020	16:22
9	L1419-06	1	Water	NA	NA	17.8	5.10	02/06/2020	16:24
10	L1430-01	1	Water	NA	NA	18.1	6.87	02/06/2020	16:25
11	L1430-01DUP	1	Water	NA	NA	18.2	6.88	02/06/2020	16:26
12	L1432-01	1	Water	NA	NA	14.5	6.53	02/06/2020	16:30
13	L1432-02	1	Water	NA	NA	14.7	5.71	02/06/2020	16:32
14	L1432-03	1	Water	NA	NA	14.9	5.85	02/06/2020	16:35
15	L1432-04	1	Water	NA	NA	15.7	5.68	02/06/2020	16:36
16	CCV2	1	Water	NA	NA	20.1	12.02	02/06/2020	16:38
17	L1432-05	1	Water	NA	NA	16.0	5.84	02/06/2020	16:40
18	L1437-02	1	Water	NA	NA	14.0	7.25	02/06/2020	16:42
19	CCV3	1	Water	NA	NA	20.1	2.01	02/06/2020	16:44

LB 107549

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W L1419 WorkList ID : 135795 Department : Wet-Chemistry Date : 02-06-2020 15:40:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2020	Water	L1419-02	pH	Cool 4 deg C	PSEG01	I61	30576	02/06/2020	9040C
02/06/2020	Water	L1419-04	pH	Cool 4 deg C	PSEG01	I61	40654	02/06/2020	9040C
02/06/2020	Water	L1419-05	pH	Cool 4 deg C	PSEG01	I61	40674	02/06/2020	9040C
02/06/2020	Water	L1419-06	pH	Cool 4 deg C	PSEG01	I61	40659	02/06/2020	9040C
02/06/2020	Water	L1430-01	pH	Cool 4 deg C	PSEG01	I61	N29305	02/06/2020	9040C
02/06/2020	Water	L1432-01	pH	Cool 4 deg C	PSEG01	I61	40646	02/06/2020	9040C
02/06/2020	Water	L1432-02	pH	Cool 4 deg C	PSEG01	I61	40647	02/06/2020	9040C
02/06/2020	Water	L1432-03	pH	Cool 4 deg C	PSEG01	I61	40648	02/06/2020	9040C
02/06/2020	Water	L1432-04	pH	Cool 4 deg C	PSEG01	I61	40649	02/06/2020	9040C
02/06/2020	Water	L1432-05	pH	Cool 4 deg C	PSEG01	I61	40650	02/06/2020	9040C
02/06/2020	Water	L1437-02	pH	Cool 4 deg C	SEAC01	L22	GROUNDWATER	02/06/2020	9040C

Date/Time 02-06-2020 15:50
 Received by: Jo
 Relinquished by: KS

Date/Time 02-06-2020 18:00
 Received by: KS
 Relinquished by: Jo