



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107601

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/06/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	7.00	02/06/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/06/2020	16:12
4	ICV	1	Water	NA	NA	20.2	7.00	02/06/2020	16:15
5	CCV1	1	Water	NA	NA	20.1	2.01	02/06/2020	16:50
6	L1419-01	1	Solid	20.03	20	15.2	8.19	02/06/2020	17:00
7	L1432-06	1	Solid	20.02	20	18.2	6.91	02/06/2020	17:02
8	L1432-06DUP	1	Solid	20.03	20	18.3	6.92	02/06/2020	17:03
9	L1432-07	1	Solid	20.04	20	17.8	7.85	02/06/2020	17:05
10	CCV2	1	Water	NA	NA	20.2	13.02	02/06/2020	17:06

WORKLIST(Hardcopy Internal Chain)

28107601

WorkList Name : PH S L1432

WorkList ID : 135796

Department : Wet-Chemistry

Date : 02-06-2020 15:41:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2020	Solid	L1419-01	pH	Cool 4 deg C	PSEG01	I61	51067	02/06/2020	9045D
02/06/2020	Solid	L1432-06	pH	Cool 4 deg C	PSEG01	I61	40657	02/06/2020	9045D
02/06/2020	Solid	L1432-07	pH	Cool 4 deg C	PSEG01	I61	40276	02/06/2020	9045D

Date/Time 02.06.2020 15:42
 Received by: JI
 Relinquished by: KS

Date/Time 02.06.2020 18:00
 Received by: KS
 Relinquished by: JA