

LB 107007


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42498



A= 0.0000e+000

B= 2.3771e-004

C= 5.6170e-002

Rho= 0.9995788

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.078	0.078	93	0.000	93				
0.2	0.200	0.210	0.010	649	0.0 %	649				
2.5	2.500	2.479	-0.021	10194	0.0 %	10194				
5.0	5.000	4.962	-0.038	20637	0.0 %	20637				
7.5	7.500	7.311	-0.189	30521	0.0 %	30521				
10.0	10.000	10.159	0.159	42498	0.0 %	42498				

LB107607 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Feb 2020 18:15:13

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.0000	7470A	PPB	06 Feb 2020 18:17:06	S	Std
0.2 - 1	30.2	649	-	0.2000	7470A	PPB	06 Feb 2020 18:19:13	S	Std
2.5 - 1	52.5	10194	-	2.5000	7470A	PPB	06 Feb 2020 18:21:21	S	Std
5.0 - 1	55	20637	-	5.0000	7470A	PPB	06 Feb 2020 18:23:28	S	Std
7.5 - 1	57.5	30521	-	7.5000	7470A	PPB	06 Feb 2020 18:25:35	S	Std
10.0 - 1	570	42498	-	10.0000	7470A	PPB	06 Feb 2020 18:27:41	S	Std
ICV05 - 1	ICV05	16089	3.8808	-	7470A	PPB	06 Feb 2020 18:30:23	U	SMPL
ICB05 - 1	ICB05	75	0.0740	-	7470A	PPB	06 Feb 2020 18:32:29	U	SMPL
CCV86 - 1	CCV86	20411	4.9081	-	7470A	PPB	06 Feb 2020 18:34:34	U	SMPL
CCB86 - 1	CCB86	90	0.0776	-	7470A	PPB	06 Feb 2020 18:36:40	U	SMPL
CRA - 1	CRA	659	0.2128	-	7470A	PPB	06 Feb 2020 18:38:45	U	SMPL
HighStd - 1	HighStd	43075	10.2957	-	7470A	PPB	06 Feb 2020 18:40:51	U	SMPL
ChkStd - 1	ChkStd	28787	6.8992	-	7470A	PPB	06 Feb 2020 18:42:57	U	SMPL
PB126620BL - 1	PBW	140	0.0894	-	7470A	PPB	06 Feb 2020 18:45:04	U	SMPL
PB126620BS - 1	LCSW	16460	3.9689	-	7470A	PPB	06 Feb 2020 18:47:11	U	SMPL
L1387-02 - 1	F06031	214	0.1070	-	7470A	PPB	06 Feb 2020 18:49:18	U	SMPL
L1387-02DUP - 1	F06031DUP	262	0.1185	-	7470A	PPB	06 Feb 2020 18:51:24	U	SMPL
L1387-02MS - 1	F06031MS	16354	3.9437	-	7470A	PPB	06 Feb 2020 18:53:32	U	SMPL
L1387-02MSD - 1	F06031MSD	16637	4.0110	-	7470A	PPB	06 Feb 2020 18:55:40	U	SMPL
L1388-02 - 1	TOTE-COMP	449	0.1629	-	7470A	PPB	06 Feb 2020 18:57:45	U	SMPL
CCV87 - 1	CCV87	21260	5.1100	-	7470A	PPB	06 Feb 2020 18:59:51	U	SMPL
CCB87 - 1	CCB87	178	0.0985	-	7470A	PPB	06 Feb 2020 19:01:56	U	SMPL
L1387-02LX5 - 1		49	0.0678	-	7470A	PPB	06 Feb 2020 19:04:02	U	SMPL
PB126620TB - 1	PB126589TB	186	0.1004	-	7470A	PPB	06 Feb 2020 19:08:14	U	SMPL
CCV88 - 1	CCV88	21088	5.0691	-	7470A	PPB	06 Feb 2020 19:10:21	U	SMPL
CCB88 - 1	CCB88	183	0.0997	-	7470A	PPB	06 Feb 2020 19:12:27	U	SMPL