



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB107612

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/07/2020	10:05
2	CAL2	1	Water	NA	NA	20.1	7.01	02/07/2020	10:06
3	CAL3	1	Water	NA	NA	20.1	10.04	02/07/2020	10:07
4	ICV	1	Water	NA	NA	20.1	6.99	02/07/2020	10:09
5	CCV1	1	Water	NA	NA	20.1	2.01	02/07/2020	10:10
6	L1440-01	1	Water	NA	NA	16.2	7.51	02/07/2020	10:11
7	L1440-01DUP	1	Water	NA	NA	16.3	7.52	02/07/2020	10:12
8	CCV2	1	Water	NA	NA	20.1	12.02	02/07/2020	10:15

LB10f612

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1440 WorkList ID : 135843 Department : Wet-Chemistry Date : 02-07-2020 09:52:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2020	Water	L1440-01	pH	Cool 4 deg C	ENTA05	L41	TW-1	02/06/2020	9040C

Date/Time 02.07.2020 10.1.00
Received by: Ja
Relinquished by: DL

Date/Time 02.07.2020 11.00
Received by: DL
Relinquished by: Ja