



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB107619

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/07/2020	09:40
2	L1419-02		1	9	100.00	>212.0	>212.0	02/07/2020	09:58
3	L1419-02DUP		1	9	100.00	>212.0	>212.0	02/07/2020	10:32
4	L1419-03		1	8	100.00	>212.0	>212.0	02/07/2020	10:59
5	L1430-02		1	7	100.00	>212.0	>212.0	02/07/2020	11:35
6	L1432-08		1	8	100.00	>212.0	>212.0	02/07/2020	12:05
7	L1437-02		1	9	100.00	>212.0	>212.0	02/07/2020	12:35
8	L1440-01		1	8	100.00	>212.0	>212.0	02/07/2020	12:56

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp2720 WorkList ID : 135829 Department : Wet-Chemistry Date : 02-07-2020 08:57:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/13/2020	Water	L1419-02	Flash Point	Cool 4 deg C	PSEG01	I61	30576	02/06/2020	1010A
02/13/2020	Water	L1419-03	Flash Point	Cool 4 deg C	PSEG01	I61	COMP-1	02/06/2020	1010A
02/13/2020	Water	L1430-02	Flash Point	Cool 4 deg C	PSEG01	I61	N29305	02/06/2020	1010A
02/13/2020	Water	L1432-08	Flash Point	Cool 4 deg C	PSEG01	I61	OD-COMP	02/06/2020	1010A
02/20/2020	Water	L1437-02	Flash Point	Cool 4 deg C	SEAC01	L22	GROUNDWATER	02/06/2020	1010A
02/07/2020	Water	L1440-01	Flash Point	Cool 4 deg C	ENTA05	L41	TW-1	02/06/2020	1010A

Date/Time 02/07/20 09:05
 Received by: KS
 Relinquished by: DK

Date/Time 02/07/20 13:05
 Received by: DK
 Relinquished by: KS