



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB107620

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	02/07/2020	12:00
2	CAL2	1	Water	NA	NA	20.1	7.01	02/07/2020	12:01
3	CAL3	1	Water	NA	NA	20.1	10.03	02/07/2020	12:04
4	ICV	1	Water	NA	NA	20.2	7.00	02/07/2020	12:05
5	CCV1	1	Water	NA	NA	20.1	2.01	02/07/2020	12:06
6	L1447-01	1	Water	NA	NA	18.6	5.05	02/07/2020	12:09
7	L1447-01DUP	1	Water	NA	NA	18.7	5.06	02/07/2020	12:10
8	CCV2	1	Water	NA	NA	20.2	12.02	02/07/2020	12:11

28107620

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH L1448 WorkList ID: 135858 Department: Wet-Chemistry Date: 02-07-2020 11:38:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2020	Water	L1447-01	pH	Cool 4 deg C	STER05	L41	POOL-WATER	02/05/2020	SM 4500-H B

Date/Time
Received by:
Relinquished by:

02/07/2020 11:40
JL
JL

Date/Time
Received by:
Relinquished by:

02/07/2020 12:35
JL
JL