



Analytical Summary Report

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB107623

ANALYST: Daniel

SUPERVISOR REVIEW BY: AMANDEEP

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP81413
IODINE SOLUTION .025N 1L	W2508
Starch Solution, 4L	W2379

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB126679BL		1	50	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	02/07/2020	12:26
2	L1430-02		1	50	50	2.00	1.92	3.82	1.90	0.10	0.02	0.16	02/07/2020	12:28
3	L1430-02DUP		1	50	50	2.00	3.82	5.70	1.88	0.12	0.04	0.32	02/07/2020	12:30
4	L1437-02		1	50	50	2.00	5.70	7.56	1.86	0.14	0.06	0.48	02/07/2020	12:32

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume