


LB 107633

7470A

INSTRUMENT ID: CY2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.050	0.050	19	0.000	19				
0.2	0.200	0.237	0.037	738	0.0 %	738				
2.5	2.500	2.259	-0.241	8517	0.0 %	8517				
5.0	5.000	5.214	0.214	19883	0.0 %	19883				
7.5	7.500	7.492	-0.008	28649	0.0 %	28649				
10.0	10.000	9.947	-0.053	38095	0.0 %	38095				

LB107633 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Feb 2020 15:15:46					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.00007470A	PPB		07 Feb 2020 15:18:20	S	Std
0.2 - 1	50.2	738	-	0.20007470A	PPB		07 Feb 2020 15:20:27	S	Std
2.5 - 1	52.5	8517	-	2.50007470A	PPB		07 Feb 2020 15:22:35	S	Std
5.0 - 1	55	19883	-	5.00007470A	PPB		07 Feb 2020 15:24:42	S	Std
7.5 - 1	57.5	28649	-	7.50007470A	PPB		07 Feb 2020 15:26:50	S	Std
10.0 - 1	510	38095	-	10.00007470A	PPB		07 Feb 2020 15:31:24	S	Std
ICV07 - 1	ICV07	15856	4.1669	-7470A	PPB		07 Feb 2020 15:36:37	U	SMPL
ICB07 - 1	ICB07	-9	0.0430	-7470A	PPB		07 Feb 2020 15:38:42	U	SMPL
CCV92 - 1	CCV92	19901	5.2183	-7470A	PPB		07 Feb 2020 15:40:48	U	SMPL
CCB92 - 1	CCB92	-8	0.0433	-7470A	PPB		07 Feb 2020 15:42:53	U	SMPL
CRA - 1	CRA	751	0.2406	-7470A	PPB		07 Feb 2020 15:44:59	U	SMPL
HighStd - 1	HighStd	37337	9.7505	-7470A	PPB		07 Feb 2020 15:47:04	U	SMPL
ChkStd - 1	ChkStd	28512	7.4566	-7470A	PPB		07 Feb 2020 15:49:10	U	SMPL
PB126684BL - 1	PBW	-30	0.0376	-7470A	PPB		07 Feb 2020 15:51:16	U	SMPL
PB126684BS - 1	LCSW	16743	4.3974	-7470A	PPB		07 Feb 2020 15:53:23	U	SMPL
L1365-08 - 1	CL-01-020520	-3	0.0446	-7470A	PPB		07 Feb 2020 15:55:30	U	SMPL
L1365-08DUP - 1	CL-01-020520DUP	32	0.0537	-7470A	PPB		07 Feb 2020 15:57:37	U	SMPL
L1365-08MS - 1	CL-01-020520MS	15792	4.1502	-7470A	PPB		07 Feb 2020 15:59:44	U	SMPL
L1365-08MSD - 1	CL-01-020520MSD	14996	3.9433	-7470A	PPB		07 Feb 2020 16:01:52	U	SMPL
L1394-09 - 1	MNR-WHITE-PLAINS-COMP	15	0.0493	-7470A	PPB		07 Feb 2020 16:03:58	U	SMPL
CCV93 - 1	CCV93	19899	5.2178	-7470A	PPB		07 Feb 2020 16:06:04	U	SMPL
CCB93 - 1	CCB93	-15	0.0415	-7470A	PPB		07 Feb 2020 16:08:10	U	SMPL
L1365-08LX5 - 1		1	0.0456	-7470A	PPB		07 Feb 2020 16:10:15	U	SMPL
L1365-08A - 1		15479	4.0689	-7470A	PPB		07 Feb 2020 16:12:22	U	SMPL
PB126684TB - 1	PB126617TB	49	0.0581	-7470A	PPB		07 Feb 2020 16:14:28	U	SMPL
CCV94 - 1	CCV94	20497	5.3732	-7470A	PPB		07 Feb 2020 16:16:34	U	SMPL
CCB94 - 1	CCB94	-8	0.0433	-7470A	PPB		07 Feb 2020 16:18:40	U	SMPL