

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

Instrument ID : Konelab

2/7/2020 15:47

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.009	0.0	0.122	
ICB1	0.018	0.0	0.011	
CCV1	1.019	0.0	0.123	
CCB1	0.025	0.0	0.012	
PB126672BL	0.022	0.0	0.011	
PB126672BS	1.032	0.0	0.125	
L1440-01	0.004	0.0	0.010	
L1440-01DUP	0.004	0.0	0.010	
L1440-01MS	0.817	0.0	0.101	
L1440-01MSD	0.815	0.0	0.100	
CCV2	1.011	0.0	0.122	
CCB2	0.023	0.0	0.012	

N	12
Mean	0.483
SD	0.4931
CV%	102.02

AK

Aquakem v. 7.2AQ1

Results from time period:

Fri Feb 07 15:15:50 2020

Fri Feb 07 15:43:27 2020

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0485	mg/l	2/7/2020 15:15:50	
0.1PPM	A	Phenolics-	P	0.1031	mg/l	2/7/2020 15:15:51	
0.25PPM	A	Phenolics-	P	0.2536	mg/l	2/7/2020 15:15:52	
0.50PPM	A	Phenolics-	P	0.5121	mg/l	2/7/2020 15:15:53	
1.0PPM	A	Phenolics-	P	0.9726	mg/l	2/7/2020 15:15:54	
2.0PPM	A	Phenolics-	P	2.0101	mg/l	2/7/2020 15:15:55	
ICV1	S	Phenolics-	P	1.0092	mg/l	2/7/2020 15:43:16	
ICB1	S	Phenolics-	P	0.0185	mg/l	2/7/2020 15:43:17	
CCV1	S	Phenolics-	P	1.0192	mg/l	2/7/2020 15:43:18	
CCB1	S	Phenolics-	P	0.0252	mg/l	2/7/2020 15:43:19	
PB126672BL	S	Phenolics-	P	0.0216	mg/l	2/7/2020 15:43:20	
PB126672BS	S	Phenolics-	P	1.0315	mg/l	2/7/2020 15:43:21	
L1440-01	S	Phenolics-	P	0.004	mg/l	2/7/2020 15:43:22	
L1440-01DUP	S	Phenolics-	P	0.0044	mg/l	2/7/2020 15:43:23	
L1440-01MS	S	Phenolics-	P	0.8168	mg/l	2/7/2020 15:43:24	
L1440-01MSD	S	Phenolics-	P	0.8147	mg/l	2/7/2020 15:43:25	
CCV2	S	Phenolics-	P	1.0112	mg/l	2/7/2020 15:43:26	
CCB2	S	Phenolics-	P	0.0231	mg/l	2/7/2020 15:43:27	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

Instrument ID : Konelab

2/7/2020 15:17


Test Phenolics-

Accepted 2/7/2020 15:17

Factor 8.919
Bias 0.009

Coeff. of det. 0.999633

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.015	0.0485	0.0500	
2	Phenol-2	0.021	0.1031	0.1000	
3	Phenol-2	0.038	0.2536	0.2500	
4	Phenol-2	0.066	0.5121	0.5000	
5	Phenol-2	0.118	0.9726	1.0000	
6	Phenol-2	0.234	2.0101	2.0000	