



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107643

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/07/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	02/07/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/07/2020	16:12
4	ICV	1	Water	NA	NA	20.1	7.00	02/07/2020	16:15
5	CCV1	1	Water	NA	NA	20.2	2.01	02/07/2020	16:16
6	L1456-02	1	Water	NA	NA	17.5	7.19	02/07/2020	16:17
7	L1456-02DUP	1	Water	NA	NA	17.6	7.20	02/07/2020	16:18
8	CCV2	1	Water	NA	NA	20.1	12.02	02/07/2020	16:20

VB107643

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1430

WorkList ID : 135874

Department : Wet-Chemistry

Date : 02-07-2020 15:43:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/07/2020	Water	L1456-02	AST Corrosivity	Cool 4 deg C	PSEG01	I62	FM-5213	02/07/2020	9040C

Date/Time 02.07.2020 16:00
 Received by: JS
 Relinquished by: JS

Date/Time 02.07.2020 18:00
 Received by: KS
 Relinquished by: JS