



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107646

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/07/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	02/07/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/07/2020	16:12
4	ICV	1	Water	NA	NA	20.1	7.00	02/07/2020	16:15
5	CCV1	1	Water	NA	NA	20.1	2.01	02/07/2020	16:25
6	L1455-01	1	Solid	20.03	20	22.0	6.42	02/07/2020	16:26
7	L1455-01DUP	1	Solid	20.04	20	22.1	6.43	02/07/2020	16:27
8	L1471-05	1	Solid	20.01	20	21.7	6.98	02/07/2020	16:30
9	CCV2	1	Water	NA	NA	20.1	12.02	02/07/2020	16:32

VB 107646

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1432 S WorkList ID : 135876 Department : Wet-Chemistry Date : 02-07-2020 15:46:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/07/2020	Solid	L1455-01	pH	Cool 4 deg C	PSEG01	I62	32193-32194	02/07/2020	9045D
02/07/2020	Solid	L1471-05	pH	Cool 4 deg C	PSEG01	L62	PPE-CLOTHING	02/07/2020	9045D

Date/Time 02.07.20 16:00
Received by: JS
Relinquished by: KS

Date/Time 02.07.2020 16:00
Received by: KS
Relinquished by: JS