



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107647

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/07/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	02/07/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/07/2020	16:12
4	ICV	1	Water	NA	NA	20.1	7.00	02/07/2020	16:15
5	CCV1	1	Water	NA	NA	20.2	2.01	02/07/2020	17:00
6	L1438-01	1	Solid	20.03	20	21.4	8.46	02/07/2020	17:02
7	L1438-02	1	Solid	20.04	20	21.4	8.25	02/07/2020	17:05
8	L1438-03	1	Solid	20.03	20	21.4	8.60	02/07/2020	17:06
9	L1456-05	1	Solid	20.03	20	21.2	7.50	02/07/2020	17:10
10	L1456-05DUP	1	Solid	20.04	20	21.3	7.51	02/07/2020	17:11
11	CCV2	1	Water	NA	NA	20.2	12.02	02/07/2020	17:12

VB 107647

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1438 WorkList ID : 135870 Department : Wet-Chemistry Date : 02-07-2020 13:45:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2020	Solid	L1438-01	Corrosivity	Cool 4 deg C	PARS15	L33	HAM-WC-S-SB-BC	02/05/2020	9045D
02/05/2020	Solid	L1438-02	Corrosivity	Cool 4 deg C	PARS15	L33	HAM-WC-S-SB-01	02/05/2020	9045D
02/05/2020	Solid	L1438-03	Corrosivity	Cool 4 deg C	PARS15	L33	HAM-WC-S-SB-0203	02/05/2020	9045D
02/07/2020	Solid	L1456-05	Corrosivity	Cool 4 deg C	PSEG01	I62	ETGI-337-290	02/07/2020	9045D

Date/Time 02-07-2020 16:00
 Received by: Jg
 Relinquished by: KS

Date/Time 02-07-2020 18:00
 Received by: KS
 Relinquished by: Jg