



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107666

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/10/2020	10:02
2	L1419-08		1	9	100.00	>212.0	>212.0	02/10/2020	10:35
3	L1419-08DUP		1	9	100.00	>212.0	>212.0	02/10/2020	10:55
4	L1432-10		1	8	100.00	>212.0	>212.0	02/10/2020	11:20
5	L1456-02		1	7	100.00	>212.0	>212.0	02/10/2020	11:50
6	L1471-01		1	8	100.00	>212.0	>212.0	02/10/2020	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB107666

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp21020 WorkList ID : 135924 Department : Wet-Chemistry Date : 02-10-2020 09:24:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/13/2020	Water	L1419-08 <i>D</i>	Flash Point	Cool 4 deg C	PSEG01	I61	51067	02/06/2020	1010A
02/13/2020	Water	L1432-10 <i>B</i>	Flash Point	Cool 4 deg C	PSEG01	I61	40276	02/06/2020	1010A
02/12/2020	Water	L1456-02 <i>C</i>	Flash Point	Cool 4 deg C	PSEG01	I62	FM-5213	02/07/2020	1010A
02/21/2020	Water	L1471-01 <i>B</i>	Flash Point	Cool 4 deg C	PSEG01	L62	30844	02/07/2020	1010A

Date/Time 02/10/20 09:30
Received by: *KS*
Relinquished by: *DL*

Date/Time 02/10/20 12:30
Received by: *DL*
Relinquished by: *KS*