

LB107669

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

Instrument ID : Konelab

2/10/2020 12:36

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.083	0.0	0.098	
ICB1	0.130	0.0	0.003	
CCV1	247.467	0.0	0.244	
CCB1	-0.023	0.0	0.003	
PB126714BL	0.057	0.0	0.003	
L1440-01	-0.326	0.0	0.002	
L1440-01DUP	-0.017	0.0	0.003	
CCV2	248.866	0.0	0.246	
CCB2	0.569	0.0	0.003	

N	9
Mean	66.090
SD	108.0998
CV%	163.57

Aquakem v. 7.2AQ1

AO

Results from time period:

Mon Feb 10 12:01:23 2020

Mon Feb 10 12:27:47 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.2947	µg/l	2/10/2020 12:01:23	
5.OPPBCN	A	Total CN	P	4.5245	µg/l	2/10/2020 12:01:24	
10PPBCN	A	Total CN	P	9.5708	µg/l	2/10/2020 12:01:25	
50PPBCN	A	Total CN	P	49.7443	µg/l	2/10/2020 12:01:26	
100PPBCN	A	Total CN	P	100.8019	µg/l	2/10/2020 12:01:27	
250PPBCN	A	Total CN	P	251.5494	µg/l	2/10/2020 12:01:28	
500PPBCN	A	Total CN	P	499.1038	µg/l	2/10/2020 12:01:29	
ICV1	S	Total CN	P	98.0825	µg/l	2/10/2020 12:27:39	
ICB1	S	Total CN	P	0.1305	µg/l	2/10/2020 12:27:40	
CCV1	S	Total CN	P	247.4666	µg/l	2/10/2020 12:27:41	
CCB1	S	Total CN	P	-0.0234	µg/l	2/10/2020 12:27:42	
PB126714BL	S	Total CN	P	0.0571	µg/l	2/10/2020 12:27:43	
L1440-01	S	Total CN	P	-0.3256	µg/l	2/10/2020 12:27:44	
L1440-01DUP	S	Total CN	P	-0.0175	µg/l	2/10/2020 12:27:45	
CCV2	S	Total CN	P	248.8664	µg/l	2/10/2020 12:27:46	
CCB2	S	Total CN	P	0.5688	µg/l	2/10/2020 12:27:47	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

Instrument ID : Konelab

2/10/2020 12:06


Test Total CN

Accepted 2/10/2020 12:06

Factor 1023
Bias 0.003

Coeff. of det. 0.999979

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.2947	0.0000	
2	5.0PPBCN	0.007	4.5245	5.0000	
3	10PPBCN	0.012	9.5708	10.0000	
4	50PPBCN	0.051	49.7443	50.0000	
5	100PPBCN	0.101	100.8019	100.0000	
6	250PPBCN	0.248	251.5494	250.0000	
7	500PPBCN	0.490	499.1038	500.0000	