



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107678

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/10/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.01	02/10/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	02/10/2020	17:32
4	ICV	1	Water	NA	NA	20.1	7.00	02/10/2020	17:35
5	CCV1	1	Water	NA	NA	20.2	2.01	02/10/2020	17:36
6	L1367-04	1	Solid	20.03	20	20.9	8.73	02/10/2020	17:37
7	L1367-04DUP	1	Solid	20.04	20	20.8	8.74	02/10/2020	17:38
8	L1372-05	1	Solid	20.02	20	21.0	10.32	02/10/2020	17:40
9	L1373-03	1	Solid	20.04	20	20.3	6.87	02/10/2020	17:42
10	L1475-09	1	Solid	20.02	20	20.2	8.32	02/10/2020	17:45
11	CCV2	1	Water	NA	NA	20.1	12.02	02/10/2020	17:46

20107678

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1367 WorkList ID : 135946 Department : Wet-Chemistry Date : 02-10-2020 17:03:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/10/2020	Solid	L1367-04	Corrosivity	Cool 4 deg C	PSEG05	I63	OK-03-021020	02/10/2020	9045D
02/10/2020	Solid	L1372-05	Corrosivity	Cool 4 deg C	PSEG05	I63	SU-01-021020	02/10/2020	9045D
02/10/2020	Solid	L1373-03	Corrosivity	Cool 4 deg C	PSEG05	I63	TR-05-021020	02/10/2020	9045D
02/10/2020	Solid	L1475-09	Corrosivity	Cool 4 deg C	PSEG01	I63	02-10-OD	02/10/2020	9045D

Date/Time 02-10-2020 17:10
 Received by: JS
 Relinquished by: JS

Date/Time 02-10-2020 18:30
 Received by: JS
 Relinquished by: JS