


LB107682

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	-19	0.000	-19				
0.2	0.200	0.183	-0.017	872	0.0 %	872				
2.5	2.500	2.461	-0.039	10455	0.0 %	10455				
5.0	5.000	5.098	0.098	21550	0.0 %	21550				
7.5	7.500	7.620	0.120	32161	0.0 %	32161				
10.0	10.000	9.866	-0.134	41610	0.0 %	41610				

LB107682 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Feb 2020 09:14:24

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7470A	PPB	11 Feb 2020 09:28:31	S	Std
0.2 - 1	50.2	872	-	0.2000	7470A	PPB	11 Feb 2020 09:30:38	S	Std
2.5 - 1	52.5	10455	-	2.5000	7470A	PPB	11 Feb 2020 09:32:45	S	Std
5.0 - 1	55	21550	-	5.0000	7470A	PPB	11 Feb 2020 09:34:53	S	Std
7.5 - 1	57.5	32161	-	7.5000	7470A	PPB	11 Feb 2020 09:36:59	S	Std
10.0 - 1	510	41610	-	10.0000	7470A	PPB	11 Feb 2020 09:39:05	S	Std
ICV10 - 1	ICV10	16214	3.8298	-	7470A	PPB	11 Feb 2020 09:41:43	U	SMPL
ICB10 - 1	ICB10	-72	-0.0413	-	7470A	PPB	11 Feb 2020 09:43:48	U	SMPL
CCV05 - 1	CCV05	20762	4.9108	-	7470A	PPB	11 Feb 2020 09:48:31	U	SMPL
CCB05 - 1	CCB05	-71	-0.0411	-	7470A	PPB	11 Feb 2020 09:50:36	U	SMPL
CRA - 1	CRA	847	0.1771	-	7470A	PPB	11 Feb 2020 09:52:42	U	SMPL
HighStd - 1	HighStd	41892	9.9334	-	7470A	PPB	11 Feb 2020 09:54:48	U	SMPL
ChkStd - 1	ChkStd	28367	6.7185	-	7470A	PPB	11 Feb 2020 09:56:54	U	SMPL
PB126717BL - 1	PBW	-64	-0.0394	-	7470A	PPB	11 Feb 2020 09:59:00	U	SMPL
PB126717BS - 1	LCSW	16426	3.8802	-	7470A	PPB	11 Feb 2020 10:03:47	U	SMPL
L1430-02 - 1	N29305	-70	-0.0409	-	7470A	PPB	11 Feb 2020 10:05:53	U	SMPL
L1430-02DUP - 1	N29305DUP	-62	-0.0390	-	7470A	PPB	11 Feb 2020 10:08:01	U	SMPL
L1430-02MS - 1	N29305MS	13828	3.2626	-	7470A	PPB	11 Feb 2020 10:10:08	U	SMPL
L1430-02MSD - 1	N29305MSD	14793	3.4920	-	7470A	PPB	11 Feb 2020 10:14:39	U	SMPL
L1454-06 - 1	INK-PIT	123	0.0050	-	7470A	PPB	11 Feb 2020 10:16:45	U	SMPL
CCV06 - 1	CCV06	20058	4.7435	-	7470A	PPB	11 Feb 2020 10:18:51	U	SMPL
CCB06 - 1	CCB06	-68	-0.0404	-	7470A	PPB	11 Feb 2020 10:20:57	U	SMPL
L1454-07 - 1	INK-TOTE	72	-0.0071	-	7470A	PPB	11 Feb 2020 10:23:02	U	SMPL
L1456-02 - 1	FM-5213	-68	-0.0404	-	7470A	PPB	11 Feb 2020 10:25:09	U	SMPL
PB126718BL - 1	PB126718BL	-19	-0.0287	-	7470A	PPB	11 Feb 2020 10:27:14	U	SMPL
PB126718BS - 1	PB126718BS	16392	3.8721	-	7470A	PPB	11 Feb 2020 10:32:37	U	SMPL
L1438-01 - 1	HAM-WC-S-SB-BC	-50	-0.0361	-	7470A	PPB	11 Feb 2020 10:34:43	U	SMPL
L1438-02 - 1	HAM-WC-S-SB-01	-42	-0.0342	-	7470A	PPB	11 Feb 2020 10:36:50	U	SMPL
L1438-03 - 1	HAM-WC-S-SB-0203	-28	-0.0309	-	7470A	PPB	11 Feb 2020 10:38:57	U	SMPL
L1456-05 - 1	ETGI-337-290	-34	-0.0323	-	7470A	PPB	11 Feb 2020 10:41:04	U	SMPL
L1456-05DUP - 1	ETGI-337-290DUP	-44	-0.0347	-	7470A	PPB	11 Feb 2020 10:43:12	U	SMPL
L1456-05MS - 1	ETGI-337-290MS	16553	3.9104	-	7470A	PPB	11 Feb 2020 10:45:18	U	SMPL
CCV07 - 1	CCV07	20305	4.8022	-	7470A	PPB	11 Feb 2020 10:47:25	U	SMPL
CCB07 - 1	CCB07	-63	-0.0392	-	7470A	PPB	11 Feb 2020 10:49:31	U	SMPL
L1456-05MSD - 1	ETGI-337-290MSD	16411	3.8766	-	7470A	PPB	11 Feb 2020 10:51:37	U	SMPL
L1471-03 - 1	30522	-20	-0.0290	-	7470A	PPB	11 Feb 2020 10:53:43	U	SMPL
L1471-05 - 1	PPE-CLOTHING	-41	-0.0340	-	7470A	PPB	11 Feb 2020 10:55:50	U	SMPL
L1430-02LX5 - 1		-47	-0.0354	-	7470A	PPB	11 Feb 2020 10:57:56	U	SMPL
L1430-02A - 1		17345	4.0986	-	7470A	PPB	11 Feb 2020 11:00:03	U	SMPL
L1456-05LX5 - 1		-49	-0.0359	-	7470A	PPB	11 Feb 2020 11:02:10	U	SMPL
L1456-05A - 1		17461	4.1262	-	7470A	PPB	11 Feb 2020 11:04:16	U	SMPL
PB126717TB - 1		-47	-0.0354	-	7470A	PPB	11 Feb 2020 11:06:24	U	SMPL
PB126718TB - 1		-29	-0.0311	-	7470A	PPB	11 Feb 2020 11:08:32	U	SMPL
CCV08 - 1	CCV08	20656	4.8856	-	7470A	PPB	11 Feb 2020 11:10:38	U	SMPL
CCB08 - 1	CCB08	-59	-0.0383	-	7470A	PPB	11 Feb 2020 11:12:45	U	SMPL