



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/12/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 13:07  
In Date: 02/11/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:30  
Out Date: 02/12/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB107684

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|--------------------------|----------------------------|---------|----------|
| L1462-01 | MGATP8          | 1      | 1.14            | 9.77                     | 8.43                       | 84.5    |          |
| L1462-02 | MGATP9          | 2      | 1.15            | 9.6                      | 7.66                       | 77.0    |          |
| L1462-03 | MGATP9D         | 3      | 1.15            | 9.6                      | 7.66                       | 77.0    |          |
| L1462-04 | MGATP9S         | 4      | 1.15            | 9.6                      | 7.66                       | 77.0    |          |
| L1462-05 | MGATQ0          | 5      | 1.14            | 9.83                     | 7.77                       | 76.3    |          |
| L1462-06 | MGATQ1          | 6      | 1.14            | 9.63                     | 7.65                       | 76.7    |          |
| L1462-07 | MGATQ2          | 7      | 1.13            | 9.94                     | 8.46                       | 83.2    |          |
| L1462-08 | MGATQ3          | 8      | 1.16            | 9.62                     | 7.65                       | 76.7    |          |
| L1462-09 | MGATQ4          | 9      | 1.13            | 9.69                     | 7.68                       | 76.5    |          |
| L1462-10 | MGATQ5          | 10     | 1.13            | 9.98                     | 7.63                       | 73.4    |          |
| L1462-11 | MGATQ6          | 11     | 1.13            | 9.53                     | 6.97                       | 69.5    |          |
| L1462-12 | MGATQ7          | 12     | 1.12            | 9.68                     | 8.33                       | 84.2    |          |
| L1462-13 | MGATQ8          | 13     | 1.13            | 9.71                     | 7.89                       | 78.8    |          |
| L1462-14 | MGATQ9          | 14     | 1.14            | 9.96                     | 8.02                       | 78.0    |          |
| L1462-15 | MGATR0          | 15     | 1.14            | 9.58                     | 7.48                       | 75.1    |          |
| L1462-16 | MGATR1          | 16     | 1.12            | 9.6                      | 7.05                       | 69.9    |          |
| L1462-17 | MGATR2          | 17     | 1.14            | 9.68                     | 7.17                       | 70.6    |          |
| L1462-18 | MGATR3          | 18     | 1.13            | 9.89                     | 7.97                       | 78.1    |          |
| L1462-19 | MGATR4          | 19     | 1.14            | 9.84                     | 9.04                       | 90.8    |          |
| L1462-20 | MGATR5          | 20     | 1.14            | 9.66                     | 7.82                       | 78.4    |          |
| L1462-21 | MGATR6          | 21     | 1.15            | 9.82                     | 7.98                       | 78.8    |          |
| L1462-22 | MGATR7          | 22     | 1.13            | 9.77                     | 7.38                       | 72.3    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$