



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107694

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/11/2020	12:35
2	CAL2	1	Water	NA	NA	20.2	7.01	02/11/2020	12:36
3	CAL3	1	Water	NA	NA	20.1	10.03	02/11/2020	12:37
4	ICV	1	Water	NA	NA	20.2	7.00	02/11/2020	12:38
5	CCV1	1	Water	NA	NA	20.2	2.01	02/11/2020	12:40
6	L1015-02	1	Water	NA	NA	15.8	7.39	02/11/2020	12:41
7	L1015-02DUP	1	Water	NA	NA	15.9	7.40	02/11/2020	12:42
8	CCV2	1	Water	NA	NA	20.1	12.02	02/11/2020	12:44

WORKLIST(Hardcopy Internal Chain)

18107694

WorkList Name : CORROSIVITY L1015 WorkList ID : 135988 Department : Wet-Chemistry Date : 02-11-2020 12:27:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2020	Water	L1015-02	-A Corrosivity	Cool 4 deg C	PSEG01	I63	PR-MANHOLE	02/11/2020	9040C

Date/Time 02.11.2020 12.30
 Received by: Jo
 Relinquished by: Jo

Date/Time 02.11.2020 13.30
 Received by: Jo
 Relinquished by: Jo