



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107695

Slope : 99.1

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2628         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 02/11/2020 | 12:35     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 02/11/2020 | 12:36     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.03       | 02/11/2020 | 12:37     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 02/11/2020 | 12:38     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 02/11/2020 | 12:40     |
| 6   | L1015-01    | 1  | Water  | NA          | NA          | 15.8             | 7.39        | 02/11/2020 | 12:41     |
| 7   | L1015-01DUP | 1  | Water  | NA          | NA          | 15.9             | 7.40        | 02/11/2020 | 12:42     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 02/11/2020 | 12:44     |

VB 107695

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1015      WorkList ID : 135987      Department : Wet-Chemistry      Date : 02-11-2020 12:27:07

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|-------------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 02/11/2020 | Water  | L1015-01-1A | pH   | Cool 4 deg C | PSEG01   | I63              | PR-MANHOLE      | 02/11/2020   | 9040C  |

Date/Time 02-11-2020 12:30  
Received by: JS  
Relinquished by: JS

Date/Time 02-11-2020 13:30  
Received by: JS  
Relinquished by: JS