



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107707

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	02/11/2020	17:25
2	CAL2	1	Water	NA	NA	20.1	7.01	02/11/2020	17:26
3	CAL3	1	Water	NA	NA	20.1	10.03	02/11/2020	17:27
4	ICV	1	Water	NA	NA	20.1	6.99	02/11/2020	17:29
5	CCV1	1	Water	NA	NA	20.1	2.01	02/11/2020	17:30
6	L1492-01	1	Solid	20.03	20	21.0	7.78	02/11/2020	17:32
7	L1492-02	1	Solid	20.04	20	20.9	8.02	02/11/2020	17:35
8	L1492-02DUP	1	Solid	20.03	20	20.8	8.03	02/11/2020	17:36
9	CCV2	1	Water	NA	NA	20.1	12.02	02/11/2020	17:38

WORKLIST(Hardcopy Internal Chain)

LB 107707

WorkList Name : PH L1492 WorkList ID : 135996 Department : Wet-Chemistry Date : 02-11-2020 16:13:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2020	Solid	L1492-01 -	pH	Cool 4 deg C	RAIL01	L61	T-12-O-(0-1.9)	02/11/2020	9045D
02/11/2020	Solid	L1492-02 -	pH	Cool 4 deg C	RAIL01	L61	T-6-E1-E4-(0-2)	02/11/2020	9045D

Date/Time 02-11-2020 16:15
 Received by: JS
 Relinquished by: KS

Date/Time 02-11-2020 17:00
 Received by: KS
 Relinquished by: JS