


LB107715

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	127	0.000	127				
0.2	0.200	0.214	0.014	1421	0.0 %	1421				
2.5	2.500	2.499	-0.001	14916	0.0 %	14916				
5.0	5.000	5.065	0.065	30079	0.0 %	30079				
7.5	7.500	7.349	-0.151	43568	0.0 %	43568				
10.0	10.000	10.077	0.077	59686	0.0 %	59686				

LB107715 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 12 Feb 2020 12:14:02					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	127	-	0.0000	7471B	PPB	12 Feb 2020 12:15:54	S	Std
0.2 - 1	50.2	1421	-	0.2000	7471B	PPB	12 Feb 2020 12:18:11	S	Std
2.5 - 1	52.5	14916	-	2.5000	7471B	PPB	12 Feb 2020 12:20:28	S	Std
5.0 - 1	55	30079	-	5.0000	7471B	PPB	12 Feb 2020 12:22:45	S	Std
7.5 - 1	57.5	43568	-	7.5000	7471B	PPB	12 Feb 2020 12:25:01	S	Std
10.0 - 1	510	59686	-	10.0000	7471B	PPB	12 Feb 2020 12:27:17	S	Std
ICV13 - 1	ICV13	22844	3.8408	-	7471B	PPB	12 Feb 2020 12:29:57	U	SMPL
ICB13 - 1	ICB13	40	-0.0194	-	7471B	PPB	12 Feb 2020 12:32:11	U	SMPL
CCV16 - 1	CCV16	30113	5.0712	-	7471B	PPB	12 Feb 2020 12:34:26	U	SMPL
CCB16 - 1	CCB16	32	-0.0208	-	7471B	PPB	12 Feb 2020 12:36:41	U	SMPL
CRA - 1	CRA	1417	0.2136	-	7471B	PPB	12 Feb 2020 12:38:57	U	SMPL
HighStd - 1	HighStd	59682	10.0766	-	7471B	PPB	12 Feb 2020 12:41:11	U	SMPL
ChkStd - 1	ChkStd	41724	7.0367	-	7471B	PPB	12 Feb 2020 12:43:27	U	SMPL
PB126758BL - 1	PBS	128	-0.0045	-	7471B	PPB	12 Feb 2020 12:45:43	U	SMPL
PB126758BS - 1	LCSS	21669	3.6419	-	7471B	PPB	12 Feb 2020 12:47:59	U	SMPL
L1367-01 - 1	OK-03-021020	1728	0.2663	-	7471B	PPB	12 Feb 2020 12:50:15	U	SMPL
L1372-01 - 1	SU-01-021020	2776	0.4437	-	7471B	PPB	12 Feb 2020 12:52:32	U	SMPL
L1373-01 - 1	TR-05-021020	2617	0.4168	-	7471B	PPB	12 Feb 2020 12:54:49	U	SMPL
L1478-01 - 1	SOIL-02	2062	0.3228	-	7471B	PPB	12 Feb 2020 12:57:06	U	SMPL
L1478-02 - 1	330	2469	0.3917	-	7471B	PPB	12 Feb 2020 12:59:21	U	SMPL
CCV17 - 1	CCV17	30221	5.0895	-	7471B	PPB	12 Feb 2020 13:01:37	U	SMPL
CCB17 - 1	CCB17	53	-0.0172	-	7471B	PPB	12 Feb 2020 13:03:52	U	SMPL
L1481-01 - 1	72-11931	1831	0.2837	-	7471B	PPB	12 Feb 2020 13:06:08	U	SMPL
L1481-01DUP - 1	72-11931DUP	1828	0.2832	-	7471B	PPB	12 Feb 2020 13:08:23	U	SMPL
L1481-01MS - 1	72-11931MS	22742	3.8235	-	7471B	PPB	12 Feb 2020 13:10:38	U	SMPL
L1481-01MSD - 1	72-11931MSD	22738	3.8228	-	7471B	PPB	12 Feb 2020 13:12:54	U	SMPL
L1481-02 - 1	72-11989	1152	0.1688	-	7471B	PPB	12 Feb 2020 13:15:09	U	SMPL
L1482-01 - 1	TP7-EP1-1.5	30330	5.1080	-	7471B	PPB	12 Feb 2020 13:17:26	U	SMPL
L1482-02 - 1	TP7-EP2-2	30958	5.2143	-	7471B	PPB	12 Feb 2020 13:19:42	U	SMPL
L1482-03 - 1	TP6-EP1-1.5	1614	0.2470	-	7471B	PPB	12 Feb 2020 13:21:59	U	SMPL
L1482-04 - 1	TP6-EP2-2	18759	3.1493	-	7471B	PPB	12 Feb 2020 13:24:16	U	SMPL
CCV18 - 1	CCV18	30319	5.1061	-	7471B	PPB	12 Feb 2020 13:26:32	U	SMPL
CCB18 - 1	CCB18	78	-0.0130	-	7471B	PPB	12 Feb 2020 13:28:47	U	SMPL
L1481-01LX5 - 1		464	0.0523	-	7471B	PPB	12 Feb 2020 13:31:02	U	SMPL
L1481-01A - 1		21304	3.5801	-	7471B	PPB	12 Feb 2020 13:33:18	U	SMPL
CCV19 - 1	CCV19	30364	5.1137	-	7471B	PPB	12 Feb 2020 13:35:33	U	SMPL
CCB19 - 1	CCB19	39	-0.0196	-	7471B	PPB	12 Feb 2020 13:37:49	U	SMPL