



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/13/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 15:00
In Date: 02/12/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:42
Out Date: 02/13/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107718

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L1506-01	DBCY1	1	1.17	9.88	7.7	75.0	
L1506-02	DBCY1MS	2	1.17	9.88	7.7	75.0	
L1506-03	DBCY1MSD	3	1.17	9.88	7.7	75.0	
L1506-04	DBCY2	4	1.16	9.86	7.88	77.2	
L1506-05	DBCY3	5	1.18	9.98	8.02	77.7	
L1506-06	DBCY4	6	1.12	9.56	7.92	80.6	
L1506-07	DBCY5	7	1.12	9.92	8.29	81.5	
L1506-08	DBCY6	8	1.13	9.96	8.3	81.2	
L1506-09	DBCY7	9	1.13	9.6	7.97	80.8	
L1506-10	DBCY8	10	1.11	9.6	8.47	86.7	
L1506-11	DBCY9	11	1.14	9.57	7.89	80.1	
L1506-12	DBCZ0	12	1.15	9.88	7.61	74.0	
L1506-13	DBCZ1	13	1.15	9.8	7.26	70.6	
L1506-14	DBCZ2	14	1.1	9.84	7.74	76.0	
L1506-15	DBCZ3	15	1.13	9.92	8.52	84.1	
L1506-16	DBCZ4	16	1.17	9.64	7.18	71.0	
L1506-17	DBCZ5	17	1.18	9.82	8.51	84.8	
L1506-18	DBCZ6	18	1.15	9.92	8.32	81.8	
L1506-19	DBCZ7	19	1.18	9.95	8.33	81.5	
L1506-20	DBCZ8	20	1.19	9.98	8.68	85.2	
L1506-21	DBCZ9	21	1.12	9.52	7.21	72.5	
L1506-22	DBD00	22	1.17	9.76	8.06	80.2	
L1506-23	DBCX9	23	1.00	2.00	2.00	100.0	P.E. SAMPLE
L1506-24	BCY0	24	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$