


LB107721

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	40	0.000	40				
0.2	0.200	0.213	0.013	815	0.0 %	815				
2.5	2.500	2.439	-0.061	10229	0.0 %	10229				
5.0	5.000	5.057	0.057	21298	0.0 %	21298				
7.5	7.500	7.408	-0.092	31240	0.0 %	31240				
10.0	10.000	10.054	0.054	42426	0.0 %	42426				

LB107721 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Feb 2020 15:11:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0000	7470A	PPB	12 Feb 2020 15:14:42	S	Std
0.2 - 1	50.2	815	-	0.2000	7470A	PPB	12 Feb 2020 15:16:49	S	Std
2.5 - 1	52.5	10229	-	2.5000	7470A	PPB	12 Feb 2020 15:18:57	S	Std
5.0 - 1	55	21298	-	5.0000	7470A	PPB	12 Feb 2020 15:21:04	S	Std
7.5 - 1	57.5	31240	-	7.5000	7470A	PPB	12 Feb 2020 15:23:11	S	Std
10.0 - 1	510	42426	-	10.0000	7470A	PPB	12 Feb 2020 15:25:17	S	Std
ICV14 - 1	ICV14	16070	3.8205	-	7470A	PPB	12 Feb 2020 15:27:50	U	SMPL
ICB14 - 1	ICB14	-7	0.0184	-	7470A	PPB	12 Feb 2020 15:29:55	U	SMPL
CCV20 - 1	CCV20	21164	5.0252	-	7470A	PPB	12 Feb 2020 15:34:32	U	SMPL
CCB20 - 1	CCB20	-8	0.0182	-	7470A	PPB	12 Feb 2020 15:36:37	U	SMPL
CRA - 1	CRA	777	0.2038	-	7470A	PPB	12 Feb 2020 15:38:42	U	SMPL
HighStd - 1	HighStd	42897	10.1649	-	7470A	PPB	12 Feb 2020 15:40:48	U	SMPL
ChkStd - 1	ChkStd	29119	6.9065	-	7470A	PPB	12 Feb 2020 15:42:54	U	SMPL
PB126776BL - 1	PBW	-25	0.0142	-	7470A	PPB	12 Feb 2020 15:45:00	U	SMPL
L1493-02 - 1	2S-DELINEATION-D2	-1	0.0198	-	7470A	PPB	12 Feb 2020 15:49:14	U	SMPL
L1493-02DUP - 1	2S-DELINEATION-D2DUP	4	0.0210	-	7470A	PPB	12 Feb 2020 15:51:21	U	SMPL
L1493-02MS - 1	2S-DELINEATION-D2MS	13674	3.2539	-	7470A	PPB	12 Feb 2020 15:53:29	U	SMPL
L1493-02MSD - 1	2S-DELINEATION-D2MSD	15576	3.7037	-	7470A	PPB	12 Feb 2020 15:55:37	U	SMPL
L1493-03 - 1	2S-DELINEATION-D1	3	0.0208	-	7470A	PPB	12 Feb 2020 15:57:42	U	SMPL
CCV21 - 1	CCV21	21028	4.9930	-	7470A	PPB	12 Feb 2020 15:59:48	U	SMPL
CCB21 - 1	CCB21	-10	0.0177	-	7470A	PPB	12 Feb 2020 16:01:54	U	SMPL
L1493-07 - 1	2S-SECTION-F	7	0.0217	-	7470A	PPB	12 Feb 2020 16:04:00	U	SMPL
L1493-10 - 1	SECTION-A	32	0.0276	-	7470A	PPB	12 Feb 2020 16:06:06	U	SMPL
L1493-11 - 1	1S-SECTION-F	15	0.0236	-	7470A	PPB	12 Feb 2020 16:08:12	U	SMPL
L1493-12 - 1	SECTION-C	24	0.0257	-	7470A	PPB	12 Feb 2020 16:10:18	U	SMPL
L1493-02LX5 - 1		26	0.0262	-	7470A	PPB	12 Feb 2020 16:12:25	U	SMPL
L1493-02A - 1		15109	3.5932	-	7470A	PPB	12 Feb 2020 16:14:32	U	SMPL
PB126776BS - 1	LCSW	15045	3.5781	-	7470A	PPB	12 Feb 2020 16:16:39	U	SMPL
PB126776TB - 1	PB126760TB	-9	0.0179	-	7470A	PPB	12 Feb 2020 16:18:46	U	SMPL
CCV22 - 1	CCV22	21339	5.0666	-	7470A	PPB	12 Feb 2020 16:20:54	U	SMPL
CCB22 - 1	CCB22	-4	0.0191	-	7470A	PPB	12 Feb 2020 16:23:00	U	SMPL