



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107727

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/12/2020	17:20
2	CAL2	1	Water	NA	NA	20.2	7.00	02/12/2020	17:22
3	CAL3	1	Water	NA	NA	20.2	10.03	02/12/2020	17:23
4	ICV	1	Water	NA	NA	20.2	7.01	02/12/2020	17:25
5	CCV1	1	Water	NA	NA	20.2	2.01	02/12/2020	17:26
6	L1370-05	1	Solid	20.03	20	20.9	8.70	02/12/2020	17:28
7	L1385-04	1	Solid	20.04	20	20.8	8.55	02/12/2020	17:29
8	L1509-11	1	Solid	20.02	20	21.0	8.21	02/12/2020	17:30
9	L1509-12	1	Solid	20.03	20	20.3	7.48	02/12/2020	17:32
10	L1509-12DUP	1	Solid	20.04	20	20.4	7.49	02/12/2020	17:33
11	CCV2	1	Water	NA	NA	20.2	12.02	02/12/2020	17:35

WORKLIST(Hardcopy Internal Chain)

107727

WorkList Name : CORROSIONITY L1509 WorkList ID : 136037 Department : Wet-Chemistry Date : 02-12-2020 15:47:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/12/2020	Solid	L1370-05	Corrosivity	Cool 4 deg C	PSEG05	I63	OR-02-021220	02/12/2020	9045D
02/12/2020	Solid	L1385-04	Corrosivity	Cool 4 deg C	PSEG05	I63	HD-02-021220	02/12/2020	9045D
02/12/2020	Solid	L1509-11	Corrosivity	Cool 4 deg C	TULL02	L63	AOC-3-SP-C	02/12/2020	9045D
02/12/2020	Solid	L1509-12	Corrosivity	Cool 4 deg C	TULL02	L63	AOC-3-SP-D	02/12/2020	9045D

Date/Time 02.12.2020 16:10
 Received by: J0
 Relinquished by: J-C

Date/Time 02.12.2020 18:10
 Received by: J-C
 Relinquished by: J0