



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107742

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/13/2020	11:55
2	L1015-02		1	8	100.00	>212.0	>212.0	02/13/2020	12:25
3	L1015-02DUP		1	8	100.00	>212.0	>212.0	02/13/2020	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 107742

WorkList Name : fp21320 WorkList ID : 136076 Department : Wet-Chemistry Date : 02-13-2020 11:23:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/24/2020	Water	L1015-02	C Flash Point	Cool 4 deg C	PSEG01	I63	PR-MANHOLE	02/11/2020	1010A

Date/Time 02/13/20 11:25
 Received by: KS
 Relinquished by: DK

Date/Time 02/13/20 12:55
 Received by: DK
 Relinquished by: KS