



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107743

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	02/13/2020	12:45
2	CAL2	1	Water	NA	NA	20.1	7.00	02/13/2020	12:46
3	CAL3	1	Water	NA	NA	20.1	10.03	02/13/2020	12:47
4	ICV	1	Water	NA	NA	20.1	7.00	02/13/2020	12:48
5	CCV1	1	Water	NA	NA	20.2	2.01	02/13/2020	12:50
6	L1522-01	1	Water	NA	NA	17.0	5.26	02/13/2020	13:00
7	L1522-01DUP	1	Water	NA	NA	17.1	5.27	02/13/2020	13:01
8	CCV2	1	Water	NA	NA	20.2	12.02	02/13/2020	13:05

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1522 WorkList ID : 136079 Department : Wet-Chemistry Date : 02-13-2020 12:38:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2020	Water	L1522-01 -17	pH	Cool 4 deg C	STER06	M11	POOL-WATER	02/11/2020	9040C

Date/Time 02-13-2020 12:40
 Received by: JP
 Relinquished by: JP

Date/Time 02-13-2020 13:40
 Received by: JP
 Relinquished by: JP