

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 02/14/2020

OVEN IN TEMP1: 98 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 02/13/2020 16:30

DATE & TIME OUT1: 02/14/2020 07:25

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 02/14/2020 08:00

DATE & TIME OUT2: 02/14/2020 09:40

Run Number: LB107763

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB107763BL	LB107763BL	54.1369	54.1369	100	54.1369	54.1369	54.1369	0.0000	0
2	L1509-13	AOC-3-SP-C	78.4807	78.4807	100	78.4996	78.4996	78.4996	0.0189	189
3	L1509-14	AOC-3-SP-D	74.7901	74.7901	100	74.8058	74.8058	74.8058	0.0157	157
4	L1509-14DUP	AOC-3-SP-DDUP	85.4472	85.4472	100	85.4629	85.4629	85.4629	0.0157	157

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VP 107763

WORKLIST(Hardcopy Internal Chain)

WorkList Name : astm ts 11509 WorkList ID : 136114 Department : Wet-Chemistry Date : 02-13-2020 15:14:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/19/2020	Solid	L1509-13	✓ ASTM TS	Cool 4 deg C	TULL02	L63	AOC-3-SP-C	02/12/2020	SM2540 B
02/19/2020	Solid	L1509-14	✓ ASTM TS	Cool 4 deg C	TULL02	L63	AOC-3-SP-D	02/12/2020	SM2540 B

Date/Time 02-13-2020 15:22
 Received by: Jc
 Relinquished by: Jc

Date/Time 02-13-2020 17:00
 Received by: Jc
 Relinquished by: Jc