


LA107767
INSTRUMENT ID: CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.064	-0.064	1766	0.000	1766				
0.2	0.200	0.147	-0.053	2897	0.0 %	2897				
2.5	2.500	2.627	0.127	16185	0.0 %	16185				
5.0	5.000	5.137	0.137	29631	0.0 %	29631				
7.5	7.500	7.336	-0.164	41411	0.0 %	41411				
10.0	10.000	10.016	0.016	55767	0.0 %	55767				

LB107767 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 14 Feb 2020 11:36:00					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1766	-	0.0000	7471B	PPB	14 Feb 2020 11:41:28	S	Std
0.2 - 1	50.2	2897	-	0.2000	7471B	PPB	14 Feb 2020 11:43:45	S	Std
2.5 - 1	52.5	16185	-	2.5000	7471B	PPB	14 Feb 2020 11:46:03	S	Std
5.0 - 1	55	29631	-	5.0000	7471B	PPB	14 Feb 2020 11:48:19	S	Std
7.5 - 1	57.5	41411	-	7.5000	7471B	PPB	14 Feb 2020 11:50:35	S	Std
10.0 - 1	510	55767	-	10.0000	7471B	PPB	14 Feb 2020 11:52:51	S	Std
ICV18 - 1	ICV18	23098	3.9178	-	7471B	PPB	14 Feb 2020 11:55:37	U	SMPL
ICB18 - 1	ICB18	1389	-0.1347	-	7471B	PPB	14 Feb 2020 11:57:51	U	SMPL
CCV32 - 1	CCV32	29406	5.0954	-	7471B	PPB	14 Feb 2020 12:00:06	U	SMPL
CCB32 - 1	CCB32	1369	-0.1384	-	7471B	PPB	14 Feb 2020 12:02:21	U	SMPL
CRA - 1	CRA	2874	0.1425	-	7471B	PPB	14 Feb 2020 12:04:36	U	SMPL
HighStd - 1	HighStd	55271	9.9237	-	7471B	PPB	14 Feb 2020 12:06:51	U	SMPL
ChkStd - 1	ChkStd	39306	6.9435	-	7471B	PPB	14 Feb 2020 12:09:06	U	SMPL
PB126808BL - 1	PBS	1200	-0.1700	-	7471B	PPB	14 Feb 2020 12:11:23	U	SMPL
PB126808BS - 1	LCSS	19748	3.2925	-	7471B	PPB	14 Feb 2020 12:13:40	U	SMPL
L1518-01 - 1	OILY-STONES	8432	1.1801	-	7471B	PPB	14 Feb 2020 12:15:56	U	SMPL
L1523-01 - 1	COMP-1	1269	-0.1571	-	7471B	PPB	14 Feb 2020 12:21:07	U	SMPL
L1523-01DUP - 1	COMP-1DUP	1323	-0.1470	-	7471B	PPB	14 Feb 2020 12:23:24	U	SMPL
L1523-01MS - 1	COMP-1MS	22425	3.7922	-	7471B	PPB	14 Feb 2020 12:25:40	U	SMPL
L1523-01MSD - 1	COMP-1MSD	22286	3.7663	-	7471B	PPB	14 Feb 2020 12:27:56	U	SMPL
CCV33 - 1	CCV33	29703	5.1508	-	7471B	PPB	14 Feb 2020 12:30:10	U	SMPL
CCB33 - 1	CCB33	1395	-0.1336	-	7471B	PPB	14 Feb 2020 12:32:25	U	SMPL
L1523-02 - 1	COMP-2	1072	-0.1939	-	7471B	PPB	14 Feb 2020 12:37:12	U	SMPL
L1523-03 - 1	COMP-3	4104	0.3721	-	7471B	PPB	14 Feb 2020 12:39:27	U	SMPL
L1528-01 - 1	FK-GF-02-048-A	3834	0.3217	-	7471B	PPB	14 Feb 2020 12:41:42	U	SMPL
L1528-03 - 1	FK-GF-02-048-B	4256	0.4005	-	7471B	PPB	14 Feb 2020 12:43:58	U	SMPL
L1529-01 - 1	FK-GF-02-049-A	5821	0.6927	-	7471B	PPB	14 Feb 2020 12:46:14	U	SMPL
L1529-03 - 1	FK-GF-02-049-B	5602	0.6518	-	7471B	PPB	14 Feb 2020 12:48:30	U	SMPL
L1523-01LX5 - 1		1615	-0.0925	-	7471B	PPB	14 Feb 2020 12:50:46	U	SMPL
L1523-01A - 1		23158	3.9290	-	7471B	PPB	14 Feb 2020 12:53:03	U	SMPL
CCV34 - 1	CCV34	29972	5.2010	-	7471B	PPB	14 Feb 2020 12:55:19	U	SMPL
CCB34 - 1	CCB34	1392	-0.1341	-	7471B	PPB	14 Feb 2020 12:57:35	U	SMPL