



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107768

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/14/2020	12:15
2	CAL2	1	Water	NA	NA	20.1	7.01	02/14/2020	12:16
3	CAL3	1	Water	NA	NA	20.1	10.03	02/14/2020	12:17
4	ICV	1	Water	NA	NA	20.2	6.99	02/14/2020	12:20
5	CCV1	1	Water	NA	NA	20.1	2.01	02/14/2020	12:21
6	L1538-01	1	Water	NA	NA	15.6	4.97	02/14/2020	12:24
7	L1538-01DUP	1	Water	NA	NA	15.7	4.98	02/14/2020	12:25
8	CCV2	1	Water	NA	NA	20.2	12.02	02/14/2020	12:30

MB 107768

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1538 WorkList ID : 136122 Department : Wet-Chemistry Date : 02-14-2020 12:09:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/13/2020	Water	L1538-01-A	pH	Cool 4 deg C	STER01	M12	POOL-WATER	02/13/2020	9040C

Date/Time 02-14-2020 12:10
 Received by: JR
 Relinquished by: JR

Date/Time 02-14-2020 16:15
 Received by: JR
 Relinquished by: JR