



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107777

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/14/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	02/14/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	02/14/2020	17:19
4	ICV	1	Water	NA	NA	20.1	7.00	02/14/2020	17:20
5	CCV1	1	Water	NA	NA	20.1	2.02	02/14/2020	17:35
6	L1548-01	1	Solid	20.03	20	20.9	6.11	02/14/2020	17:36
7	L1548-01DUP	1	Solid	20.04	20	20.8	6.12	02/14/2020	17:37
8	CCV2	1	Water	NA	NA	20.1	12.02	02/14/2020	17:38

28108777

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1548 WorkList ID : 136131 Department : Wet-Chemistry Date : 02-14-2020 15:50:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/14/2020	Solid	L1548-01	pH	Cool 4 deg C	TULL02	M13	FK-PS-01-063	02/14/2020	9045D

Date/Time 02-14-2020 16:00
Received by: JS
Relinquished by: JS

Date/Time 02-14-2020 16:35
Received by: JS
Relinquished by: JS