


LB107793

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3033e-004

C= -3.4828e-003

Rho= 0.9984978

Accept=Accepted

Slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.00	0.000	0.002	0.002	25	0.000	25					-
0.05	0.050	0.055	0.005	252	0.0 %	252					10
0.20	0.200	0.218	0.018	960	0.0 %	960					9
2.50	2.500	2.219	-0.281	9651	0.0 %	9651					-11
5.00	5.000	5.357	0.357	23275	0.0 %	23275					7
7.50	7.500	7.646	0.146	33213	0.0 %	33213					2
10.0	10.000	9.752	-0.248	42356	0.0 %	42356					-2

LB107793
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0	S0.0	25	-	0	ISM02.4	PPB	2/17/2020 14:14	Std
0.05	S0.05	252	-	0.05	ISM02.4	PPB	2/17/2020 14:16	Std
0.2	S0.2	960	-	0.2	ISM02.4	PPB	2/17/2020 14:18	Std
2.5	S2.5	9651	-	2.5	ISM02.4	PPB	2/17/2020 14:21	Std
5	S5	23275	-	5	ISM02.4	PPB	2/17/2020 14:23	Std
7.5	S7.5	33213	-	7.5	ISM02.4	PPB	2/17/2020 14:25	Std
10	S10.0	42356	-	10	ISM02.4	PPB	2/17/2020 14:36	Std
ICV	ICV	18224	4.194 -		ISM02.4	PPB	2/17/2020 14:39	SMPL
ICB	ICB	1	-0.0033 -		ISM02.4	PPB	2/17/2020 14:41	SMPL
CCV041	CCV041	23199	5.3399 -		ISM02.4	PPB	2/17/2020 14:44	SMPL
CCB041	CCB041	-16	-0.0072 -		ISM02.4	PPB	2/17/2020 14:46	SMPL
PB126864BL	PBW864	9	-0.0014 -		ISM02.4	PPB	2/17/2020 14:48	SMPL
L1519-01	METC40	31	0.0037 -		ISM02.4	PPB	2/17/2020 14:50	SMPL
L1519-02	METC41	5	-0.0023 -		ISM02.4	PPB	2/17/2020 14:53	SMPL
L1519-03	METC42	12	-0.0007 -		ISM02.4	PPB	2/17/2020 14:55	SMPL
L1519-04	METC42D	6	-0.0021 -		ISM02.4	PPB	2/17/2020 14:57	SMPL
L1519-05	METC42S	4987	1.1452 -		ISM02.4	PPB	2/17/2020 14:59	SMPL
CCV042	CCV042	23464	5.4009 -		ISM02.4	PPB	2/17/2020 15:02	SMPL
CCB042	CCB042	5	-0.0023 -		ISM02.4	PPB	2/17/2020 15:04	SMPL