

LB157794

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : Dk

Instrument ID : Konelab

2/17/2020 14:13

Test: ^{Reactive} Total CN Dk 0.912
avg

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.682	0.0	0.096	
ICB1	0.268	0.0	0.003	
CCV1	237.189	0.0	0.235	
CCB1	0.127	0.0	0.003	
PB126853BL	0.096	0.0	0.002	
L1365-16	0.054	0.0	0.002	
L1365-16DUP	0.209	0.0	0.003	
L1544-03	0.102	0.0	0.002	
L1544-07	0.349	0.0	0.003	
CCV2	239.254	0.0	0.237	
CCB2	0.571	0.0	0.003	

N	11
Mean	52.173
SD	96.2884
CV%	184.56

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 17 14:07:12 2020

Mon Feb 17 14:07:22 2020

Sample Id	Sam/Ctr/c	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.2356	µg/l	2/17/2020 11:03:26	
5.OPPBCN	A	Reactive CN	P	4.8219	µg/l	2/17/2020 11:03:27	
10PPBCN	A	Reactive CN	P	9.8834	µg/l	2/17/2020 11:03:28	
50PPBCN	A	Reactive CN	P	50.2433	µg/l	2/17/2020 11:03:29	
100PPBCN	A	Reactive CN	P	100.1324	µg/l	2/17/2020 11:03:30	
250PPBCN	A	Reactive CN	P	250.4026	µg/l	2/17/2020 11:03:31	
500PPBCN	A	Reactive CN	P	499.752	µg/l	2/17/2020 11:03:32	
ICV1	S	Reactive CN	P	95.6821	µg/l	2/17/2020 14:07:12	
ICB1	S	Reactive CN	P	0.268	µg/l	2/17/2020 14:07:13	
CCV1	S	Reactive CN	P	237.1888	µg/l	2/17/2020 14:07:14	
CCB1	S	Reactive CN	P	0.1265	µg/l	2/17/2020 14:07:15	
PB126853E	S	Reactive CN	P	0.0961	µg/l	2/17/2020 14:07:16	
L1365-16	S	Reactive CN	P	0.0539	µg/l	2/17/2020 14:07:17	
L1365-16D	S	Reactive CN	P	0.2093	µg/l	2/17/2020 14:07:18	
L1544-03	S	Reactive CN	P	0.1019	µg/l	2/17/2020 14:07:19	
L1544-07	S	Reactive CN	P	0.349	µg/l	2/17/2020 14:07:20	
CCV2	S	Reactive CN	P	239.2537	µg/l	2/17/2020 14:07:21	
CCB2	S	Reactive CN	P	0.5713	µg/l	2/17/2020 14:07:22	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DK

Instrument ID : Konelab

2/17/2020 11:19


Test ~~Total~~ ^{Reactive} CN

Accepted 2/17/2020 11:18

Factor 1020
Bias 0.002

Coeff. of det. 0.999998

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.2356	0.0000	
2	5.0PPBCN	0.007	4.8219	5.0000	
3	10PPBCN	0.012	9.8834	10.0000	
4	50PPBCN	0.052	50.2433	50.0000	
5	100PPBCN	0.101	100.1324	100.0000	
6	250PPBCN	0.248	250.4026	250.0000	
7	500PPBCN	0.492	499.7520	500.0000	