



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107795

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	02/17/2020	14:40
2	CAL2	1	Water	NA	NA	21.2	7.01	02/17/2020	14:42
3	CAL3	1	Water	NA	NA	21.2	10.04	02/17/2020	14:43
4	ICV	1	Water	NA	NA	20.9	7.02	02/17/2020	14:46
5	CCV1	1	Water	NA	NA	21.1	2.01	02/17/2020	14:48
6	L1562-01	1	Water	NA	NA	20.7	5.17	02/17/2020	14:50
7	L1562-01DUP	1	Water	NA	NA	20.7	5.18	02/17/2020	14:52
8	CCV2	1	Water	NA	NA	21.1	12.02	02/17/2020	14:54

LB107795

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHwater-2172020 WorkList ID : 136173 Department : Wet-Chemistry Date : 02-17-2020 13:51:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/13/2020	Water	L1562-01	pH	Cool 4 deg C	STER04	M13	POOL-WATER	02/13/2020	SM 4500-H B

Date/Time 2/17/2020 14:35
Received by: AM
Relinquished by: CS

Date/Time 2/17/2020 15:00
Received by: AR
Relinquished by: AM