



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB107812

ANALYST: ketankumar
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include various calibration standards and reagents like Hexachrome 1 PPM, Hexachrome 0.5 PPM, etc.

Intercept: 0.0007 Slope: 0.7548 Regression: 0.999994

Main data table with columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: ketankumar

Parameter: Hexavalent Chromiu

SUPERVISOR REVIEW BY: apatel

Run Number: LB107812

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.04	0.000	0.389	0.389	0.514	02/18/2020	11:57
2	ICB		1	100	100		2.03	0.000	0.000	0.000	-0.001	02/18/2020	11:58
3	CCV1	0.5	1	100	100		2.06	0.000	0.381	0.381	0.504	02/18/2020	12:05
4	CCB1		1	100	100		2.05	0.000	0.000	0.000	-0.001	02/18/2020	12:06
5	lb107812BL		1	100	100		2.04	0.000	0.000	0.000	-0.001	02/18/2020	12:07
6	lb107812BS	0.5	1	100	100		2.08	0.000	0.408	0.408	0.540	02/18/2020	12:08
7	L1566-01		1	100	100		2.03	0.001	0.001	0.000	-0.001	02/18/2020	12:09
8	L1566-01DUP		1	100	100		2.04	0.000	0.000	0.000	-0.001	02/18/2020	12:10
9	L1566-01MS	1	2	100	100		2.05	0.000	0.394	0.394	0.521	02/18/2020	12:11
10	L1566-01MSD	1	2	100	100		2.06	0.000	0.393	0.393	0.520	02/18/2020	12:12
11	L1566-02		1	100	100		2.04	0.003	0.003	0.000	-0.001	02/18/2020	12:13
12	CCV2	0.5	1	100	100		2.06	0.000	0.382	0.382	0.505	02/18/2020	12:14
13	CCB2		1	100	100		2.07	0.000	0.000	0.000	-0.001	02/18/2020	12:15

LB107812

WORKLIST(Hardcopy Internal Chain)

WorkList Name : hex21820-1 WorkList ID : 136206 Department : Wet-Chemistry Date : 02-18-2020 10:26:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/15/2020	Water	L1566-01	Hexavalent Chromium	Cool 4 deg C	AQUA04	M41	0119-FIELD-BLANK	02/14/2020	7196A
02/15/2020	Water	L1566-02	Hexavalent Chromium	Cool 4 deg C	AQUA04	M41	0130-SITE-WATER	02/14/2020	7196A

Date/Time 02/18/20 11:40
 Received by: KS
 Relinquished by: DK

Date/Time 02/18/20 12:20
 Received by: DK
 Relinquished by: KS